

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021022\
 Data File : PP043586.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Feb 2022 22:56
 Operator : YP\AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP021022AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 11 02:45:26 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 11 02:43:45 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.920	3.175	112.0E6	97614585	50.082	49.051
2)	SA Decachlor...	10.072	8.555	62947997	83198469	49.668	48.576
Target Compounds							
21)	L5 AR-1248-1	5.174	4.322	20680409	18153633	504.430	487.343
22)	L5 AR-1248-2	5.472	4.576	28342094	23940576	507.587	489.368
23)	L5 AR-1248-3	5.691	4.619	33190639	25174863	497.489	490.593
24)	L5 AR-1248-4	6.131	4.801	34356101	28975499	491.736	494.243
25)	L5 AR-1248-5	6.172	5.222	33726933	28066241	487.870	489.833



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021022\
Data File : PP043586.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Feb 2022 22:56
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP021022AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 11 02:45:26 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 11 02:43:45 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.920 min
Delta R.T.: 0.000 min
Response: 111967655
Conc: 50.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP021022AR1248




#1 Tetrachloro-m-xylene

R.T.: 3.175 min
Delta R.T.: 0.001 min
Response: 97614585
Conc: 49.05 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.072 min
Delta R.T.: 0.002 min
Response: 62947997
Conc: 49.67 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.555 min
Delta R.T.: 0.000 min
Response: 83198469
Conc: 48.58 ng/ml



#21 AR-1248-1


R.T.: 5.174 min
Delta R.T.: 0.000 min
Response: 20680409
Conc: 504.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP021022AR1248




#21 AR-1248-1

R.T.: 4.322 min
Delta R.T.: 0.000 min
Response: 18153633
Conc: 487.34 ng/ml




#22 AR-1248-2

R.T.: 5.472 min
Delta R.T.: 0.001 min
Response: 28342094
Conc: 507.59 ng/ml



#22 AR-1248-2


R.T.: 4.576 min
Delta R.T.: 0.000 min
Response: 23940576
Conc: 489.37 ng/ml



#23 AR-1248-3


R.T.: 5.691 min
Delta R.T.: 0.000 min
Response: 33190639
Conc: 497.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP021022AR1248




#23 AR-1248-3

R.T.: 4.619 min
Delta R.T.: 0.000 min
Response: 25174863
Conc: 490.59 ng/ml




#24 AR-1248-4

R.T.: 6.131 min
Delta R.T.: 0.001 min
Response: 34356101
Conc: 491.74 ng/ml



#24 AR-1248-4


R.T.: 4.801 min
Delta R.T.: 0.000 min
Response: 28975499
Conc: 494.24 ng/ml



#25 AR-1248-5

R.T.: 6.172 min
Delta R.T.: 0.001 min
Response: 33726933
Conc: 487.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP021022AR1248



#25 AR-1248-5

R.T.: 5.222 min
Delta R.T.: 0.000 min
Response: 28066241
Conc: 489.83 ng/ml