

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021023\
 Data File : PP055538.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Feb 2023 16:57
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 10 21:54:10 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.328	3.573	84588685	76646855	48.650	51.057
2) SA Decachlor...	10.083	8.583	80820139	64984870	46.672	48.117
Target Compounds						
21) L5 AR-1248-1	5.498	4.655	19195126	15797410	460.274	509.542
22) L5 AR-1248-2	5.774	4.892	28382065	23150509	446.440	506.271
23) L5 AR-1248-3	5.978	4.934	30746060	24384643	444.576	511.495
24) L5 AR-1248-4	6.384	5.106	33307947	28885174	439.222	523.025
25) L5 AR-1248-5	6.422	5.499	32406783	25072108	439.158	509.897



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021023\
Data File : PP055538.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Feb 2023 16:57
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 10 21:54:10 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.328 min
Delta R.T.: -0.004 min
Response: 84588685
Conc: 48.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.573 min
Delta R.T.: -0.005 min
Response: 76646855
Conc: 51.06 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.083 min
Delta R.T.: -0.008 min
Response: 80820139
Conc: 46.67 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.583 min
Delta R.T.: -0.015 min
Response: 64984870
Conc: 48.12 ng/ml



#21 AR-1248-1


R.T.: 5.498 min
Delta R.T.: -0.004 min
Response: 19195126
Conc: 460.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.655 min
Delta R.T.: -0.008 min
Response: 15797410
Conc: 509.54 ng/ml




#22 AR-1248-2

R.T.: 5.774 min
Delta R.T.: -0.004 min
Response: 28382065
Conc: 446.44 ng/ml



#22 AR-1248-2


R.T.: 4.892 min
Delta R.T.: -0.009 min
Response: 23150509
Conc: 506.27 ng/ml



#23 AR-1248-3


R.T.: 5.978 min
Delta R.T.: -0.004 min
Response: 30746060
Conc: 444.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.934 min
Delta R.T.: -0.009 min
Response: 24384643
Conc: 511.50 ng/ml




#24 AR-1248-4

R.T.: 6.384 min
Delta R.T.: -0.004 min
Response: 33307947
Conc: 439.22 ng/ml



#24 AR-1248-4


R.T.: 5.106 min
Delta R.T.: -0.008 min
Response: 28885174
Conc: 523.02 ng/ml



#25 AR-1248-5

R.T.: 6.422 min
Delta R.T.: -0.004 min
Response: 32406783
Conc: 439.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.499 min
Delta R.T.: -0.009 min
Response: 25072108
Conc: 509.90 ng/ml