

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021122\
 Data File : PP043593.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Feb 2022 09:49
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 12 04:38:19 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 11 04:48:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.919	3.174	112.6E6	97779201	50.910	52.058
2) SA Decachlor...	10.072	8.556	64705399	84567004	54.471	52.125
Target Compounds						
26) L6 AR-1254-1	6.101	5.179	32544909	46119227	430.736	528.087
27) L6 AR-1254-2	6.342	5.339	56049026	38541834	499.686	505.286
28) L6 AR-1254-3	6.742	5.772	59713593	60922017	516.051	497.086
29) L6 AR-1254-4	7.056	6.022	41050394	39855712	514.200	504.982
30) L6 AR-1254-5	7.517	6.474	42104334	58225752	490.938	503.489



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021122\
Data File : PP043593.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Feb 2022 09:49
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 12 04:38:19 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 11 04:48:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.919 min
Delta R.T.: 0.000 min
Response: 112565439
Conc: 50.91 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.174 min
Delta R.T.: 0.000 min
Response: 97779201
Conc: 52.06 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.072 min
Delta R.T.: 0.000 min
Response: 64705399
Conc: 54.47 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.556 min
Delta R.T.: 0.001 min
Response: 84567004
Conc: 52.13 ng/ml



#26 AR-1254-1


R.T.: 6.101 min
Delta R.T.: -0.001 min
Response: 32544909
Conc: 430.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.179 min
Delta R.T.: 0.000 min
Response: 46119227
Conc: 528.09 ng/ml




#27 AR-1254-2

R.T.: 6.342 min
Delta R.T.: 0.000 min
Response: 56049026
Conc: 499.69 ng/ml



#27 AR-1254-2


R.T.: 5.339 min
Delta R.T.: 0.000 min
Response: 38541834
Conc: 505.29 ng/ml



#28 AR-1254-3


R.T.: 6.742 min
Delta R.T.: 0.000 min
Response: 59713593
Conc: 516.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.772 min
Delta R.T.: 0.000 min
Response: 60922017
Conc: 497.09 ng/ml




#29 AR-1254-4

R.T.: 7.056 min
Delta R.T.: 0.000 min
Response: 41050394
Conc: 514.20 ng/ml



#29 AR-1254-4


R.T.: 6.022 min
Delta R.T.: 0.000 min
Response: 39855712
Conc: 504.98 ng/ml



#30 AR-1254-5

R.T.: 7.517 min
Delta R.T.: 0.000 min
Response: 42104334
Conc: 490.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 58225752
Conc: 503.49 ng/ml