

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021122\
 Data File : PP043613.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Feb 2022 16:32
 Operator : YP\AJ
 Sample : N1487-01
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 251488

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/14/2022
 Supervised By :Ankita Jodhani 02/14/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 12 04:46:23 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 11 04:48:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.918	3.173	25268561	29075923	11.428m	15.480m#
2) SA Decachlor...	10.068	8.554	16268870	23329598	13.696	14.380

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021122\
Data File : PP043613.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Feb 2022 16:32
Operator : YP\AJ
Sample : N1487-01
Misc :
ALS Vial : 19 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
251488


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/14/2022
Supervised By :Ankita Jodhani 02/14/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 12 04:46:23 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 11 04:48:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.918 min
Delta R.T.: -0.001 min
Response: 25268561
Conc: 11.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
251488


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/14/2022
Supervised By :Ankita Jodhani 02/14/2022




#1 Tetrachloro-m-xylene

R.T.: 3.173 min
Delta R.T.: -0.001 min
Response: 29075923
Conc: 15.48 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.068 min
Delta R.T.: -0.003 min
Response: 16268870
Conc: 13.70 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.554 min
Delta R.T.: 0.000 min
Response: 23329598
Conc: 14.38 ng/ml