

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021225\
 Data File : PP069673.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Feb 2025 10:27
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 12 22:23:09 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012825.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 29 00:13:21 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.531	3.837	75930822	48291550	53.044	54.106
2) SA Decachlor...	10.259	8.897	51740908	48212097	47.343	43.105
Target Compounds						
26) L6 AR-1254-1	6.532	5.735	27683414	28541956	478.424	518.419
27) L6 AR-1254-2	6.749	5.882	38866695	25508390	484.369	527.373
28) L6 AR-1254-3	7.111	6.288	37827411	39213397	448.785	516.672
29) L6 AR-1254-4	7.394	6.516	30634139	26370466	449.261	514.548
30) L6 AR-1254-5	7.810	6.935	30906522	32374569	454.199	486.394



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021225\
Data File : PP069673.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Feb 2025 10:27
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 12 22:23:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012825.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 29 00:13:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.531 min
Delta R.T.: -0.007 min
Response: 75930822
Conc: 53.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.837 min
Delta R.T.: -0.003 min
Response: 48291550
Conc: 54.11 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.259 min
Delta R.T.: -0.020 min
Response: 51740908
Conc: 47.34 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.897 min
Delta R.T.: -0.014 min
Response: 48212097
Conc: 43.10 ng/ml



#26 AR-1254-1


R.T.: 6.532 min
Delta R.T.: -0.008 min
Response: 27683414
Conc: 478.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.735 min
Delta R.T.: -0.007 min
Response: 28541956
Conc: 518.42 ng/ml




#27 AR-1254-2

R.T.: 6.749 min
Delta R.T.: -0.008 min
Response: 38866695
Conc: 484.37 ng/ml



#27 AR-1254-2


R.T.: 5.882 min
Delta R.T.: -0.007 min
Response: 25508390
Conc: 527.37 ng/ml



#28 AR-1254-3


R.T.: 7.111 min
Delta R.T.: -0.009 min
Response: 37827411
Conc: 448.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.288 min
Delta R.T.: -0.007 min
Response: 39213397
Conc: 516.67 ng/ml




#29 AR-1254-4

R.T.: 7.394 min
Delta R.T.: -0.009 min
Response: 30634139
Conc: 449.26 ng/ml



#29 AR-1254-4


R.T.: 6.516 min
Delta R.T.: -0.007 min
Response: 26370466
Conc: 514.55 ng/ml



#30 AR-1254-5

R.T.: 7.810 min
Delta R.T.: -0.010 min
Response: 30906522
Conc: 454.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.935 min
Delta R.T.: -0.009 min
Response: 32374569
Conc: 486.39 ng/ml