

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021323\
 Data File : PP055565.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Feb 2023 10:03
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 13 20:08:48 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.328	3.575	84924178	73628656	48.843	49.046
2) SA Decachlor...	10.084	8.584	81271835	65641404	46.933	48.603
Target Compounds						
21) L5 AR-1248-1	5.498	4.656	19181621	14919834	459.950	481.236
22) L5 AR-1248-2	5.773	4.892	28121170	22459974	442.336	491.170
23) L5 AR-1248-3	5.978	4.934	30525380	23498637	441.385	492.910
24) L5 AR-1248-4	6.383	5.105	32375755	27339670	426.929	495.040
25) L5 AR-1248-5	6.422	5.499	31142355	23761322	422.023	483.240



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021323\
Data File : PP055565.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Feb 2023 10:03
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 13 20:08:48 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.328 min
Delta R.T.: -0.004 min
Response: 84924178
Conc: 48.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.575 min
Delta R.T.: -0.003 min
Response: 73628656
Conc: 49.05 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.084 min
Delta R.T.: -0.007 min
Response: 81271835
Conc: 46.93 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.584 min
Delta R.T.: -0.014 min
Response: 65641404
Conc: 48.60 ng/ml



#21 AR-1248-1


R.T.: 5.498 min
Delta R.T.: -0.003 min
Response: 19181621
Conc: 459.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.656 min
Delta R.T.: -0.008 min
Response: 14919834
Conc: 481.24 ng/ml




#22 AR-1248-2

R.T.: 5.773 min
Delta R.T.: -0.004 min
Response: 28121170
Conc: 442.34 ng/ml



#22 AR-1248-2


R.T.: 4.892 min
Delta R.T.: -0.009 min
Response: 22459974
Conc: 491.17 ng/ml



#23 AR-1248-3


R.T.: 5.978 min
Delta R.T.: -0.004 min
Response: 30525380
Conc: 441.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.934 min
Delta R.T.: -0.009 min
Response: 23498637
Conc: 492.91 ng/ml




#24 AR-1248-4

R.T.: 6.383 min
Delta R.T.: -0.004 min
Response: 32375755
Conc: 426.93 ng/ml



#24 AR-1248-4


R.T.: 5.105 min
Delta R.T.: -0.009 min
Response: 27339670
Conc: 495.04 ng/ml



#25 AR-1248-5

R.T.: 6.422 min
Delta R.T.: -0.004 min
Response: 31142355
Conc: 422.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.499 min
Delta R.T.: -0.009 min
Response: 23761322
Conc: 483.24 ng/ml