

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021323\
 Data File : PP055596.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Feb 2023 22:21
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 14 04:05:22 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.329	3.575	92433547	82732317	53.162	55.110
2) SA Decachlor...	10.085	8.584	82684707	65337018	47.748	48.378
Target Compounds						
26) L6 AR-1254-1	6.357	5.458	34974686	42508994	443.321	526.374
27) L6 AR-1254-2	6.577	5.606	53040205	37067253	438.265	517.970
28) L6 AR-1254-3	6.945	6.011	54446262	58302207	444.496	523.324
29) L6 AR-1254-4	7.232	6.239	38825533	30175783	440.638	490.604
30) L6 AR-1254-5	7.653	6.659	44395638	47077522	438.862	496.577



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021323\
Data File : PP055596.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Feb 2023 22:21
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 14 04:05:22 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.329 min
Delta R.T.: -0.003 min
Response: 92433547
Conc: 53.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.575 min
Delta R.T.: -0.003 min
Response: 82732317
Conc: 55.11 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.085 min
Delta R.T.: -0.006 min
Response: 82684707
Conc: 47.75 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.584 min
Delta R.T.: -0.014 min
Response: 65337018
Conc: 48.38 ng/ml



#26 AR-1254-1


R.T.: 6.357 min
Delta R.T.: -0.004 min
Response: 34974686
Conc: 443.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.458 min
Delta R.T.: -0.009 min
Response: 42508994
Conc: 526.37 ng/ml




#27 AR-1254-2

R.T.: 6.577 min
Delta R.T.: -0.004 min
Response: 53040205
Conc: 438.27 ng/ml



#27 AR-1254-2


R.T.: 5.606 min
Delta R.T.: -0.009 min
Response: 37067253
Conc: 517.97 ng/ml



#28 AR-1254-3


R.T.: 6.945 min
Delta R.T.: -0.004 min
Response: 54446262
Conc: 444.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.011 min
Delta R.T.: -0.009 min
Response: 58302207
Conc: 523.32 ng/ml




#29 AR-1254-4

R.T.: 7.232 min
Delta R.T.: -0.004 min
Response: 38825533
Conc: 440.64 ng/ml



#29 AR-1254-4


R.T.: 6.239 min
Delta R.T.: -0.010 min
Response: 30175783
Conc: 490.60 ng/ml



#30 AR-1254-5

R.T.: 7.653 min
Delta R.T.: -0.004 min
Response: 44395638
Conc: 438.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.659 min
Delta R.T.: -0.010 min
Response: 47077522
Conc: 496.58 ng/ml