

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021325\
 Data File : PP069706.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Feb 2025 09:32
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 13 12:50:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012825.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 29 00:13:21 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.539	3.839	78666469	48380343	54.955	54.205
2) SA Decachlor...	10.274	8.900	60205163	51962515	55.088	46.458
Target Compounds						
16) L4 AR-1242-1	5.692	4.929	21414967	14761340	513.483	511.977
17) L4 AR-1242-2	5.713	4.948	31539093	20643359	534.643	523.973
18) L4 AR-1242-3	5.776	5.126	19484577	11504828	525.210	511.233
19) L4 AR-1242-4	5.873	5.210	16022042	11012178	488.556	510.243
20) L4 AR-1242-5	6.604	5.736	16673271	14583323	511.696	506.922



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021325\
Data File : PP069706.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Feb 2025 09:32
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 13 12:50:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012825.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 29 00:13:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.539 min
Delta R.T.: 0.000 min
Response: 78666469
Conc: 54.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.839 min
Delta R.T.: -0.001 min
Response: 48380343
Conc: 54.21 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.274 min
Delta R.T.: -0.004 min
Response: 60205163
Conc: 55.09 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.900 min
Delta R.T.: -0.011 min
Response: 51962515
Conc: 46.46 ng/ml



#16 AR-1242-1


R.T.: 5.692 min
Delta R.T.: 0.000 min
Response: 21414967
Conc: 513.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.929 min
Delta R.T.: -0.004 min
Response: 14761340
Conc: 511.98 ng/ml




#17 AR-1242-2

R.T.: 5.713 min
Delta R.T.: 0.000 min
Response: 31539093
Conc: 534.64 ng/ml



#17 AR-1242-2


R.T.: 4.948 min
Delta R.T.: -0.004 min
Response: 20643359
Conc: 523.97 ng/ml



#18 AR-1242-3


R.T.: 5.776 min
Delta R.T.: 0.000 min
Response: 19484577
Conc: 525.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.126 min
Delta R.T.: -0.003 min
Response: 11504828
Conc: 511.23 ng/ml




#19 AR-1242-4

R.T.: 5.873 min
Delta R.T.: 0.000 min
Response: 16022042
Conc: 488.56 ng/ml



#19 AR-1242-4


R.T.: 5.210 min
Delta R.T.: -0.004 min
Response: 11012178
Conc: 510.24 ng/ml



#20 AR-1242-5

R.T.: 6.604 min
Delta R.T.: 0.000 min
Response: 16673271
Conc: 511.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.736 min
Delta R.T.: -0.005 min
Response: 14583323
Conc: 506.92 ng/ml