

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021422\
 Data File : PP043683.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Feb 2022 15:12
 Operator : YP\AJ
 Sample : N1521-02MS
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TP1-AMS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 15 00:33:06 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 11 04:48:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.919	3.174	48544381	40504071	21.955	21.565
2) SA Decachlor...	10.069	8.553	25605973	34757597	21.556	21.424
Target Compounds						
3) L1 AR-1016-1	5.173	4.322	34544823	29843600	476.072	487.089
4) L1 AR-1016-2	5.196	4.340	49602006	41716378	470.292	493.749
5) L1 AR-1016-3	5.263	4.526	30983855	23282622	470.557	498.377
6) L1 AR-1016-4	5.368	4.575	25940858	18221928	478.346	487.596
7) L1 AR-1016-5	5.689	4.800	24601435	22918604	468.350	501.692
31) L7 AR-1260-1	6.920	5.911	42095545	41446426	502.680	518.275
32) L7 AR-1260-2	7.204	6.118	46330024	49777185	511.817	509.736
33) L7 AR-1260-3	7.599	6.281	30321046	47800322	418.287	523.602 #
34) L7 AR-1260-4	7.846	6.795	37235556	36835153	433.696	455.260
35) L7 AR-1260-5	8.190	7.061	65623194	85241086	423.356	461.099



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021422\
Data File : PP043683.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Feb 2022 15:12
Operator : YP\AJ
Sample : N1521-02MS
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP1-AMS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 15 00:33:06 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 11 04:48:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.919 min
Delta R.T.: 0.000 min
Response: 48544381
Conc: 21.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP1-AMS




#1 Tetrachloro-m-xylene

R.T.: 3.174 min
Delta R.T.: 0.000 min
Response: 40504071
Conc: 21.56 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.069 min
Delta R.T.: -0.003 min
Response: 25605973
Conc: 21.56 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.553 min
Delta R.T.: -0.002 min
Response: 34757597
Conc: 21.42 ng/ml



#3 AR-1016-1


R.T.: 5.173 min
Delta R.T.: 0.000 min
Response: 34544823
Conc: 476.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP1-AMS




#3 AR-1016-1

R.T.: 4.322 min
Delta R.T.: -0.001 min
Response: 29843600
Conc: 487.09 ng/ml




#4 AR-1016-2

R.T.: 5.196 min
Delta R.T.: -0.001 min
Response: 49602006
Conc: 470.29 ng/ml



#4 AR-1016-2


R.T.: 4.340 min
Delta R.T.: -0.001 min
Response: 41716378
Conc: 493.75 ng/ml



#5 AR-1016-3


R.T.: 5.263 min
Delta R.T.: 0.000 min
Response: 30983855
Conc: 470.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP1-AMS




#5 AR-1016-3

R.T.: 4.526 min
Delta R.T.: -0.002 min
Response: 23282622
Conc: 498.38 ng/ml




#6 AR-1016-4

R.T.: 5.368 min
Delta R.T.: -0.002 min
Response: 25940858
Conc: 478.35 ng/ml



#6 AR-1016-4


R.T.: 4.575 min
Delta R.T.: 0.000 min
Response: 18221928
Conc: 487.60 ng/ml



#7 AR-1016-5


R.T.: 5.689 min
Delta R.T.: -0.002 min
Response: 24601435
Conc: 468.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP1-AMS




#7 AR-1016-5

R.T.: 4.800 min
Delta R.T.: -0.001 min
Response: 22918604
Conc: 501.69 ng/ml




#31 AR-1260-1

R.T.: 6.920 min
Delta R.T.: -0.003 min
Response: 42095545
Conc: 502.68 ng/ml



#31 AR-1260-1


R.T.: 5.911 min
Delta R.T.: -0.002 min
Response: 41446426
Conc: 518.28 ng/ml



#32 AR-1260-2


R.T.: 7.204 min
Delta R.T.: -0.002 min
Response: 46330024
Conc: 511.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP1-AMS




#32 AR-1260-2

R.T.: 6.118 min
Delta R.T.: -0.002 min
Response: 49777185
Conc: 509.74 ng/ml




#33 AR-1260-3

R.T.: 7.599 min
Delta R.T.: -0.002 min
Response: 30321046
Conc: 418.29 ng/ml



#33 AR-1260-3


R.T.: 6.281 min
Delta R.T.: -0.002 min
Response: 47800322
Conc: 523.60 ng/ml



#34 AR-1260-4


R.T.: 7.846 min
Delta R.T.: -0.003 min
Response: 37235556
Conc: 433.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP1-AMS




#34 AR-1260-4

R.T.: 6.795 min
Delta R.T.: -0.002 min
Response: 36835153
Conc: 455.26 ng/ml



#35 AR-1260-5

R.T.: 8.190 min
Delta R.T.: -0.002 min
Response: 65623194
Conc: 423.36 ng/ml



#35 AR-1260-5

R.T.: 7.061 min
Delta R.T.: -0.002 min
Response: 85241086
Conc: 461.10 ng/ml