

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021423\
 Data File : PP055625.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Feb 2023 11:19
 Operator : YP\AJ
 Sample : 01529-01DL 4X
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SB-06-0-2.0DL

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 02/16/2023
 Supervised By :Sohil Jodhani 02/16/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 14 20:31:50 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.325	3.572	8115661	6511144	4.668	4.337
2) SA Decachlor...	10.075	8.579	13672887	8079769	7.896	5.983m
Target Compounds						
26) L6 AR-1254-1	6.354	5.455	23931229	26692288	303.340	330.521
27) L6 AR-1254-2	6.574	5.603	36319601	27855100	300.105	389.241 #
28) L6 AR-1254-3	6.942	6.007	50259607	54364691	410.316	487.980
29) L6 AR-1254-4	7.228	6.237	28657308	18333093	325.237	298.063
30) L6 AR-1254-5	7.650	6.655	48094401	87566175	475.425	923.655 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021423\
Data File : PP055625.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Feb 2023 11:19
Operator : YP\AJ
Sample : 01529-01DL 4X
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
SB-06-0-2.0DL


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 02/16/2023
Supervised By :Sohil Jodhani 02/16/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 14 20:31:50 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.325 min
Delta R.T.: -0.007 min
Response: 8115661
Conc: 4.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-06-0-2.0DL


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 02/16/2023
Supervised By :Sohil Jodhani 02/16/2023




#1 Tetrachloro-m-xylene

R.T.: 3.572 min
Delta R.T.: -0.007 min
Response: 6511144
Conc: 4.34 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.075 min
Delta R.T.: -0.015 min
Response: 13672887
Conc: 7.90 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.579 min
Delta R.T.: -0.020 min
Response: 8079769
Conc: 5.98 ng/ml m




#26 AR-1254-1

R.T.: 6.354 min
Delta R.T.: -0.007 min
Response: 23931229
Conc: 303.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-06-0-2.0DL


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 02/16/2023
Supervised By :Sohil Jodhani 02/16/2023




#26 AR-1254-1

R.T.: 5.455 min
Delta R.T.: -0.012 min
Response: 26692288
Conc: 330.52 ng/ml




#27 AR-1254-2

R.T.: 6.574 min
Delta R.T.: -0.008 min
Response: 36319601
Conc: 300.10 ng/ml



#27 AR-1254-2

R.T.: 5.603 min
Delta R.T.: -0.012 min
Response: 27855100
Conc: 389.24 ng/ml




#28 AR-1254-3

R.T.: 6.942 min
Delta R.T.: -0.007 min
Response: 50259607
Conc: 410.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-06-0-2.0DL


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 02/16/2023
Supervised By :Sohil Jodhani 02/16/2023




#28 AR-1254-3

R.T.: 6.007 min
Delta R.T.: -0.012 min
Response: 54364691
Conc: 487.98 ng/ml




#29 AR-1254-4

R.T.: 7.228 min
Delta R.T.: -0.008 min
Response: 28657308
Conc: 325.24 ng/ml



#29 AR-1254-4

R.T.: 6.237 min
Delta R.T.: -0.013 min
Response: 18333093
Conc: 298.06 ng/ml




#30 AR-1254-5

R.T.: 7.650 min
Delta R.T.: -0.008 min
Response: 48094401
Conc: 475.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-06-0-2.0DL

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 02/16/2023
Supervised By :Sohil Jodhani 02/16/2023



#30 AR-1254-5

R.T.: 6.655 min
Delta R.T.: -0.014 min
Response: 87566175
Conc: 923.65 ng/ml