

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021423\
 Data File : PP055637.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Feb 2023 15:12
 Operator : YP\AJ
 Sample : 01529-08DL 4X
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SB-05-5.0-7.0DL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 14 20:44:15 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.326	3.572	8060054	6695286	4.636	4.460
2) SA Decachlor...	10.078	8.580	15890132	12638218	9.176	9.358
Target Compounds						
26) L6 AR-1254-1	6.355	5.456	29077716	32201786	368.574	398.744
27) L6 AR-1254-2	6.575	5.604	42896819	31235040	354.452	436.471
28) L6 AR-1254-3	6.943	6.009	47031458	58652279	383.962	526.466 #
29) L6 AR-1254-4	7.230	6.238	25789185	20743294	292.686	337.249
30) L6 AR-1254-5	7.650	6.657	57279089	49143448	566.218	518.369



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021423\
Data File : PP055637.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Feb 2023 15:12
Operator : YP\AJ
Sample : 01529-08DL 4X
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SB-05-5.0-7.0DL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 14 20:44:15 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.326 min
Delta R.T.: -0.006 min
Response: 8060054
Conc: 4.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-05-5.0-7.0DL




#1 Tetrachloro-m-xylene

R.T.: 3.572 min
Delta R.T.: -0.006 min
Response: 6695286
Conc: 4.46 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.078 min
Delta R.T.: -0.013 min
Response: 15890132
Conc: 9.18 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.580 min
Delta R.T.: -0.019 min
Response: 12638218
Conc: 9.36 ng/ml



#26 AR-1254-1


R.T.: 6.355 min
Delta R.T.: -0.006 min
Response: 29077716
Conc: 368.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-05-5.0-7.0DL




#26 AR-1254-1

R.T.: 5.456 min
Delta R.T.: -0.011 min
Response: 32201786
Conc: 398.74 ng/ml




#27 AR-1254-2

R.T.: 6.575 min
Delta R.T.: -0.006 min
Response: 42896819
Conc: 354.45 ng/ml



#27 AR-1254-2


R.T.: 5.604 min
Delta R.T.: -0.011 min
Response: 31235040
Conc: 436.47 ng/ml



#28 AR-1254-3


R.T.: 6.943 min
Delta R.T.: -0.006 min
Response: 47031458
Conc: 383.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-05-5.0-7.0DL




#28 AR-1254-3

R.T.: 6.009 min
Delta R.T.: -0.011 min
Response: 58652279
Conc: 526.47 ng/ml




#29 AR-1254-4

R.T.: 7.230 min
Delta R.T.: -0.007 min
Response: 25789185
Conc: 292.69 ng/ml



#29 AR-1254-4


R.T.: 6.238 min
Delta R.T.: -0.012 min
Response: 20743294
Conc: 337.25 ng/ml



#30 AR-1254-5

R.T.: 7.650 min
Delta R.T.: -0.008 min
Response: 57279089
Conc: 566.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB-05-5.0-7.0DL



#30 AR-1254-5

R.T.: 6.657 min
Delta R.T.: -0.012 min
Response: 49143448
Conc: 518.37 ng/ml