

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP021522\  
 Data File : PP043758.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 15 Feb 2022 18:59  
 Operator : YP\AJ  
 Sample : N1544-07 10X  
 Misc :  
 ALS Vial : 25 Sample Multiplier: 1

Instrument :  
 ECD\_P  
 ClientSampleId :  
 07-A-07-B-07-C

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Feb 16 00:45:12 2022  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP021022.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Fri Feb 11 04:48:05 2022  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1   | RT#2  | Resp#1  | Resp#2  | ng/ml  | ng/ml   |
|-----------------------------|--------|-------|---------|---------|--------|---------|
| -----                       |        |       |         |         |        |         |
| System Monitoring Compounds |        |       |         |         |        |         |
| 1) SA Tetrachlo...          | 3.918  | 3.175 | 3334504 | 4098727 | 1.508m | 2.182 # |
| 2) SA Decachlor...          | 10.064 | 8.549 | 2765591 | 3486672 | 2.328m | 2.149   |

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP021522\  
Data File : PP043758.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 15 Feb 2022 18:59  
Operator : YP\AJ  
Sample : N1544-07 10X  
Misc :  
ALS Vial : 25 Sample Multiplier: 1

Instrument :  
ECD\_P  
ClientSampleId :  
07-A-07-B-07-C

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Feb 16 00:45:12 2022  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP021022.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Fri Feb 11 04:48:05 2022  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.918 min  
Delta R.T.: -0.001 min  
Response: 3334504  
Conc: 1.51 ng/ml

Instrument :  
ECD\_P  
ClientSampleId :  
07-A-07-B-07-C




#1 Tetrachloro-m-xylene

R.T.: 3.175 min  
Delta R.T.: 0.000 min  
Response: 4098727  
Conc: 2.18 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.064 min  
Delta R.T.: -0.008 min  
Response: 2765591  
Conc: 2.33 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.549 min  
Delta R.T.: -0.006 min  
Response: 3486672  
Conc: 2.15 ng/ml