

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021524\
 Data File : PP063309.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Feb 2024 10:20
 Operator : YP\AJ
 Sample : P1467-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 01A-01B-01C

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/16/2024
 Supervised By :Ankita Jodhani 02/16/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 16 02:04:43 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020124.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 05 12:13:11 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.492	3.645	50555964	44922231	21.136	18.900
2) SA Decachlor...	10.360	8.690	52103133	58846692	23.169	23.756m
Target Compounds						
26) L6 AR-1254-1	6.536	5.545	38794501	38646861	434.737m	319.536 #
27) L6 AR-1254-2	6.756	5.694	36638779	30289185	277.830m	280.216
28) L6 AR-1254-3	7.125	6.100	38999189	49654905	283.696m	293.601
29) L6 AR-1254-4	7.413	6.329	28900265	29119390	351.712m	364.886
30) L6 AR-1254-5	7.835	6.750	53010620	72760868	514.943m	516.465

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021524\
Data File : PP063309.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Feb 2024 10:20
Operator : YP\AJ
Sample : P1467-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
01A-01B-01C

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/16/2024
Supervised By :Ankita Jodhani 02/16/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 16 02:04:43 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020124.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 05 12:13:11 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

