

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021623\
 Data File : PP055734.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Feb 2023 16:09
 Operator : YP\AJ
 Sample : 01546-01
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 33910

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 02/17/2023
 Supervised By :Sohil Jodhani 02/17/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 16 20:39:27 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 03:37:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.329	3.575	13448243	12679134	7.735	8.446
2) SA Decachlor...	10.083	8.581	11685441	10178979	6.748	7.537

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021623\
Data File : PP055734.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Feb 2023 16:09
Operator : YP\AJ
Sample : 01546-01
Misc :
ALS Vial : 18 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
33910


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 02/17/2023
Supervised By :Sohil Jodhani 02/17/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 16 20:39:27 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP012423.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 03:37:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.329 min
Delta R.T.: -0.003 min
Response: 13448243
Conc: 7.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
33910


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 02/17/2023
Supervised By :Sohil Jodhani 02/17/2023




#1 Tetrachloro-m-xylene

R.T.: 3.575 min
Delta R.T.: -0.003 min
Response: 12679134
Conc: 8.45 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.083 min
Delta R.T.: -0.008 min
Response: 11685441
Conc: 6.75 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.581 min
Delta R.T.: -0.017 min
Response: 10178979
Conc: 7.54 ng/ml