

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021723\
 Data File : PP055772.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Feb 2023 23:47
 Operator : YP\AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 18 07:57:51 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Feb 18 07:54:13 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.324	3.568	117.1E6	104.0E6	74.672	75.255
2) SA Decachlor...	10.072	8.574	151.5E6	82559368	72.787	71.651
Target Compounds						
26) L6 AR-1254-1	6.354	5.454	51664094	53122773	701.990	732.665
27) L6 AR-1254-2	6.574	5.602	81277903	45757042	706.654	731.800
28) L6 AR-1254-3	6.941	6.006	89609271	72075969	720.995	738.563
29) L6 AR-1254-4	7.228	6.235	67349289	40730076	724.947	736.677
30) L6 AR-1254-5	7.649	6.654	77578104	59163250	734.074	724.929



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021723\
Data File : PP055772.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Feb 2023 23:47
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 18 07:57:51 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Feb 18 07:54:13 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.324 min
Delta R.T.: 0.000 min
Response: 117117126
Conc: 74.67 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.568 min
Delta R.T.: 0.000 min
Response: 103992138
Conc: 75.25 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.072 min
Delta R.T.: -0.001 min
Response: 151524134
Conc: 72.79 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.574 min
Delta R.T.: 0.000 min
Response: 82559368
Conc: 71.65 ng/ml



#26 AR-1254-1


R.T.: 6.354 min
Delta R.T.: 0.000 min
Response: 51664094
Conc: 701.99 ng/ml

Instrument :
ECD_P
ClientSampleId :




#26 AR-1254-1

R.T.: 5.454 min
Delta R.T.: 0.001 min
Response: 53122773
Conc: 732.67 ng/ml




#27 AR-1254-2

R.T.: 6.574 min
Delta R.T.: 0.000 min
Response: 81277903
Conc: 706.65 ng/ml



#27 AR-1254-2


R.T.: 5.602 min
Delta R.T.: 0.000 min
Response: 45757042
Conc: 731.80 ng/ml



#28 AR-1254-3


R.T.: 6.941 min
Delta R.T.: 0.000 min
Response: 89609271
Conc: 720.99 ng/ml

Instrument :
ECD_P
ClientSampleId :




#28 AR-1254-3

R.T.: 6.006 min
Delta R.T.: 0.001 min
Response: 72075969
Conc: 738.56 ng/ml




#29 AR-1254-4

R.T.: 7.228 min
Delta R.T.: 0.000 min
Response: 67349289
Conc: 724.95 ng/ml



#29 AR-1254-4


R.T.: 6.235 min
Delta R.T.: 0.000 min
Response: 40730076
Conc: 736.68 ng/ml



#30 AR-1254-5

R.T.: 7.649 min
Delta R.T.: 0.000 min
Response: 77578104
Conc: 734.07 ng/ml

Instrument :
ECD_P
ClientSampleId :



#30 AR-1254-5

R.T.: 6.654 min
Delta R.T.: 0.000 min
Response: 59163250
Conc: 724.93 ng/ml