

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021723\
 Data File : PP055783.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Feb 2023 02:47
 Operator : YP\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 20 04:41:40 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 20 04:36:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.324	3.567	80534116	70865600	46.694	45.702
2) SA Decachlor...	10.072	8.572	103.2E6	56504991	50.146	48.855
Target Compounds						
16) L4 AR-1242-1	5.495	4.650	24484624	17739762	484.420	494.616
17) L4 AR-1242-2	5.517	4.668	35770861	25302930	481.945	455.513
18) L4 AR-1242-3	5.579	4.845	22889026	13809270	488.020	458.444
19) L4 AR-1242-4	5.678	4.928	18489948	14953647	487.443	465.792
20) L4 AR-1242-5	6.418	5.452	21347654	16586137	496.104	471.630



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021723\
Data File : PP055783.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Feb 2023 02:47
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 20 04:41:40 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 20 04:36:53 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.324 min
Delta R.T.: -0.001 min
Response: 80534116
Conc: 46.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.567 min
Delta R.T.: 0.000 min
Response: 70865600
Conc: 45.70 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.072 min
Delta R.T.: -0.003 min
Response: 103163500
Conc: 50.15 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.572 min
Delta R.T.: -0.003 min
Response: 56504991
Conc: 48.85 ng/ml



#16 AR-1242-1


R.T.: 5.495 min
Delta R.T.: -0.001 min
Response: 24484624
Conc: 484.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#16 AR-1242-1

R.T.: 4.650 min
Delta R.T.: 0.000 min
Response: 17739762
Conc: 494.62 ng/ml




#17 AR-1242-2

R.T.: 5.517 min
Delta R.T.: -0.001 min
Response: 35770861
Conc: 481.95 ng/ml



#17 AR-1242-2


R.T.: 4.668 min
Delta R.T.: -0.001 min
Response: 25302930
Conc: 455.51 ng/ml



#18 AR-1242-3


R.T.: 5.579 min
Delta R.T.: -0.002 min
Response: 22889026
Conc: 488.02 ng/ml

Instrument :
ECD_P
ClientSampleId :




#18 AR-1242-3

R.T.: 4.845 min
Delta R.T.: -0.001 min
Response: 13809270
Conc: 458.44 ng/ml




#19 AR-1242-4

R.T.: 5.678 min
Delta R.T.: -0.001 min
Response: 18489948
Conc: 487.44 ng/ml



#19 AR-1242-4


R.T.: 4.928 min
Delta R.T.: -0.002 min
Response: 14953647
Conc: 465.79 ng/ml



#20 AR-1242-5

R.T.: 6.418 min
Delta R.T.: -0.002 min
Response: 21347654
Conc: 496.10 ng/ml

Instrument :
ECD_P
ClientSampleId :



#20 AR-1242-5

R.T.: 5.452 min
Delta R.T.: -0.002 min
Response: 16586137
Conc: 471.63 ng/ml