

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021723\
 Data File : PP055781.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Feb 2023 02:14
 Operator : YP\AJ
 Sample : AR1268ICC050
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 19 04:45:12 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Feb 19 04:45:03 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.323	3.567	6860981	6004103	4.361	4.283
2) SA Decachlor...	10.072	8.574	17413693	9441801	4.748	4.762
Target Compounds						
41) L9 AR-1268-1	8.575	7.478	12901644	8027170	45.206	47.742
42) L9 AR-1268-2	8.669	7.543	12438811	7293710	46.652	47.684
43) L9 AR-1268-3	8.896	7.748	11335834	6440347	48.169	46.937
44) L9 AR-1268-4	9.322	8.038	4184003	2714344	42.379	49.336
45) L9 AR-1268-5	9.734	8.324	34276645	18371638	45.530	45.547



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021723\
Data File : PP055781.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Feb 2023 02:14
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 19 04:45:12 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Feb 19 04:45:03 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.323 min
Delta R.T.: 0.000 min
Response: 6860981
Conc: 4.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.567 min
Delta R.T.: 0.000 min
Response: 6004103
Conc: 4.28 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.072 min
Delta R.T.: 0.000 min
Response: 17413693
Conc: 4.75 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.574 min
Delta R.T.: 0.000 min
Response: 9441801
Conc: 4.76 ng/ml



#41 AR-1268-1


R.T.: 8.575 min
Delta R.T.: 0.000 min
Response: 12901644
Conc: 45.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050




#41 AR-1268-1

R.T.: 7.478 min
Delta R.T.: 0.000 min
Response: 8027170
Conc: 47.74 ng/ml




#42 AR-1268-2

R.T.: 8.669 min
Delta R.T.: 0.000 min
Response: 12438811
Conc: 46.65 ng/ml



#42 AR-1268-2


R.T.: 7.543 min
Delta R.T.: 0.000 min
Response: 7293710
Conc: 47.68 ng/ml



#43 AR-1268-3


R.T.: 8.896 min
Delta R.T.: 0.000 min
Response: 11335834
Conc: 48.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050




#43 AR-1268-3

R.T.: 7.748 min
Delta R.T.: 0.000 min
Response: 6440347
Conc: 46.94 ng/ml




#44 AR-1268-4

R.T.: 9.322 min
Delta R.T.: 0.000 min
Response: 4184003
Conc: 42.38 ng/ml



#44 AR-1268-4


R.T.: 8.038 min
Delta R.T.: 0.000 min
Response: 2714344
Conc: 49.34 ng/ml



#45 AR-1268-5

R.T.: 9.734 min
Delta R.T.: 0.000 min
Response: 34276645
Conc: 45.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050



#45 AR-1268-5

R.T.: 8.324 min
Delta R.T.: 0.000 min
Response: 18371638
Conc: 45.55 ng/ml