

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021821\
 Data File : PP033409.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Feb 2021 19:44
 Operator : DD\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 19 00:36:24 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 09 04:00:57 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.725	3.997	3210260	2149251	52.351	52.559
2) SA Decachlor...	10.538	9.256	2280844	1777717	52.360	53.066
Target Compounds						
26) L6 AR-1254-1	6.925	6.106	896628	939685	477.559	491.686
27) L6 AR-1254-2	7.152	6.265	1453878	799006	495.907	487.310
28) L6 AR-1254-3	7.530	6.684	1465970	1292248	482.656	480.804
29) L6 AR-1254-4	7.824	6.923	1183822	775048	511.752	519.090
30) L6 AR-1254-5	8.251	7.349	1190209	1169998	491.866	483.195



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021821\
Data File : PP033409.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Feb 2021 19:44
Operator : DD\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 19 00:36:24 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020821.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 09 04:00:57 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.725 min
Delta R.T.: -0.005 min
Response: 3210260
Conc: 52.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.997 min
Delta R.T.: -0.006 min
Response: 2149251
Conc: 52.56 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.538 min
Delta R.T.: -0.013 min
Response: 2280844
Conc: 52.36 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.256 min
Delta R.T.: -0.012 min
Response: 1777717
Conc: 53.07 ng/ml



#26 AR-1254-1


R.T.: 6.925 min
Delta R.T.: -0.006 min
Response: 896628
Conc: 477.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.106 min
Delta R.T.: -0.009 min
Response: 939685
Conc: 491.69 ng/ml




#27 AR-1254-2

R.T.: 7.152 min
Delta R.T.: -0.006 min
Response: 1453878
Conc: 495.91 ng/ml



#27 AR-1254-2


R.T.: 6.265 min
Delta R.T.: -0.009 min
Response: 799006
Conc: 487.31 ng/ml



#28 AR-1254-3


R.T.: 7.530 min
Delta R.T.: -0.006 min
Response: 1465970
Conc: 482.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.684 min
Delta R.T.: -0.009 min
Response: 1292248
Conc: 480.80 ng/ml




#29 AR-1254-4

R.T.: 7.824 min
Delta R.T.: -0.007 min
Response: 1183822
Conc: 511.75 ng/ml



#29 AR-1254-4


R.T.: 6.923 min
Delta R.T.: -0.010 min
Response: 775048
Conc: 519.09 ng/ml



#30 AR-1254-5

R.T.: 8.251 min
Delta R.T.: -0.007 min
Response: 1190209
Conc: 491.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.349 min
Delta R.T.: -0.009 min
Response: 1169998
Conc: 483.19 ng/ml