

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021821\
 Data File : PP033416.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Feb 2021 21:41
 Operator : DD\AJ
 Sample : M1440-01
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SP-227-233-A

Manual Integrations
 APPROVED

Ankita
 2/19/2021 9:34:30 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 19 00:40:05 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 09 04:00:57 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.725	3.998	1541717	1039079	25.141	25.410
2) SA Decachlor...	10.539	9.255	961127	969003	22.064	28.925m#

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021821\
Data File : PP033416.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Feb 2021 21:41
Operator : DD\AJ
Sample : M1440-01
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
SP-227-233-A


Manual Integrations
APPROVED

Ankita
2/19/2021 9:34:30 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 19 00:40:05 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020821.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 09 04:00:57 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.725 min
Delta R.T.: -0.005 min
Response: 1541717
Conc: 25.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
SP-227-233-A


Manual Integrations
APPROVED

Ankita
2/19/2021 9:34:30 AM




#1 Tetrachloro-m-xylene

R.T.: 3.998 min
Delta R.T.: -0.006 min
Response: 1039079
Conc: 25.41 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.539 min
Delta R.T.: -0.012 min
Response: 961127
Conc: 22.06 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.255 min
Delta R.T.: -0.013 min
Response: 969003
Conc: 28.93 ng/ml m