

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021921\
 Data File : PP033466.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Feb 2021 1:42
 Operator : DD\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 20 03:35:04 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 09 04:00:57 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.723	3.995	1438490	902345	23.458	22.067
2)	SA Decachlor...	10.541	9.255	1000376	752424	22.965	22.460
Target Compounds							
9)	L2 AR-1221-2	5.069	0.000	26076	0	56.053	N.D. #
20)	L4 AR-1242-5	7.024f	6.129	124107	59279	119.454	70.204 #
24)	L5 AR-1248-4	6.973	0.000	627293	0	364.569	N.D. #
25)	L5 AR-1248-5	7.024f	0.000	124107	0	71.326	N.D. #
26)	L6 AR-1254-1	6.938	6.129	567173	59279	302.085	31.017 #
27)	L6 AR-1254-2	7.152	0.000	665620	0	227.038	N.D. #
28)	L6 AR-1254-3	7.543	0.000	2565565	0	844.687	N.D. #
33)	L7 AR-1260-3	8.327	0.000	353187	0	137.997	N.D. #
34)	L7 AR-1260-4	8.533f	0.000	541730	0	221.742	N.D. #
36)	L8 AR-1262-1	8.327	0.000	353187	0	92.902	N.D. #
43)	L9 AR-1268-3	9.457	0.000	196159	0	31.960	N.D. #
45)	L9 AR-1268-5	10.236	0.000	24895	0	1.521	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP021921\
Data File : PP033466.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Feb 2021 1:42
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 20 03:35:04 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020821.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 09 04:00:57 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.723 min
Delta R.T.: -0.007 min
Response: 1438490
Conc: 23.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.995 min
Delta R.T.: -0.008 min
Response: 902345
Conc: 22.07 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.541 min
Delta R.T.: -0.010 min
Response: 1000376
Conc: 22.96 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.255 min
Delta R.T.: -0.014 min
Response: 752424
Conc: 22.46 ng/ml



#9 AR-1221-2


R.T.: 5.069 min
Delta R.T.: -0.004 min
Response: 26076
Conc: 56.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 4.356 min
Response: 0
Conc: N.D.




#20 AR-1242-5

R.T.: 7.024 min
Delta R.T.: 0.025 min
Response: 124107
Conc: 119.45 ng/ml



#20 AR-1242-5


R.T.: 6.129 min
Delta R.T.: 0.013 min
Response: 59279
Conc: 70.20 ng/ml



#24 AR-1248-4


R.T.: 6.973 min
Delta R.T.: 0.012 min
Response: 627293
Conc: 364.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 5.738 min
Response: 0
Conc: N.D.




#25 AR-1248-5

R.T.: 7.024 min
Delta R.T.: 0.025 min
Response: 124107
Conc: 71.33 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T. : 6.160 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 6.938 min
Delta R.T.: 0.008 min
Response: 567173
Conc: 302.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#26 AR-1254-1

R.T.: 6.129 min
Delta R.T.: 0.014 min
Response: 59279
Conc: 31.02 ng/ml




#27 AR-1254-2

R.T.: 7.152 min
Delta R.T.: -0.007 min
Response: 665620
Conc: 227.04 ng/ml



#27 AR-1254-2


R.T.: 0.000 min
Exp R.T. : 6.274 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 7.543 min
Delta R.T.: 0.006 min
Response: 2565565
Conc: 844.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.693 min
Response: 0
Conc: N.D.




#33 AR-1260-3

R.T.: 8.327 min
Delta R.T.: -0.006 min
Response: 353187
Conc: 138.00 ng/ml



#33 AR-1260-3


R.T.: 0.000 min
Exp R.T. : 7.179 min
Response: 0
Conc: N.D.



#34 AR-1260-4


R.T.: 8.533 min
Delta R.T.: -0.029 min
Response: 541730
Conc: 221.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#34 AR-1260-4

R.T.: 0.000 min
Exp R.T. : 7.659 min
Response: 0
Conc: N.D.




#36 AR-1262-1

R.T.: 8.327 min
Delta R.T.: -0.005 min
Response: 353187
Conc: 92.90 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T. : 7.357 min
Response: 0
Conc: N.D.



#43 AR-1268-3


R.T.: 9.457 min
Delta R.T.: -0.006 min
Response: 196159
Conc: 31.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 8.459 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 10.236 min
Delta R.T.: -0.010 min
Response: 24895
Conc: 1.52 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T. : 9.031 min
Response: 0
Conc: N.D.