

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022023\
 Data File : PP055789.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Feb 2023 09:30
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 21 00:41:38 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 20 05:24:03 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.324	3.567	82630155	72561403	51.284	49.694
2)	SA Decachlor...	10.077	8.575	102.7E6	55643692	50.335	49.047
Target Compounds							
16)	L4 AR-1242-1	5.495	4.650	24770980	17857336	513.002	485.345
17)	L4 AR-1242-2	5.518	4.668	36431329	26528614	512.692	503.659
18)	L4 AR-1242-3	5.580	4.844	23278678	14192839	521.777	501.327
19)	L4 AR-1242-4	5.679	4.929	18731235	15269722	518.433	494.921
20)	L4 AR-1242-5	6.419	5.452	21241679	16563758	507.092	497.271



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022023\
Data File : PP055789.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Feb 2023 09:30
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 21 00:41:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 20 05:24:03 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.324 min
Delta R.T.: -0.001 min
Response: 82630155
Conc: 51.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.567 min
Delta R.T.: -0.002 min
Response: 72561403
Conc: 49.69 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.077 min
Delta R.T.: 0.001 min
Response: 102681429
Conc: 50.34 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.575 min
Delta R.T.: -0.002 min
Response: 55643692
Conc: 49.05 ng/ml



#16 AR-1242-1


R.T.: 5.495 min
Delta R.T.: -0.002 min
Response: 24770980
Conc: 513.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.650 min
Delta R.T.: -0.002 min
Response: 17857336
Conc: 485.35 ng/ml




#17 AR-1242-2

R.T.: 5.518 min
Delta R.T.: -0.002 min
Response: 36431329
Conc: 512.69 ng/ml



#17 AR-1242-2


R.T.: 4.668 min
Delta R.T.: -0.002 min
Response: 26528614
Conc: 503.66 ng/ml



#18 AR-1242-3


R.T.: 5.580 min
Delta R.T.: -0.001 min
Response: 23278678
Conc: 521.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.844 min
Delta R.T.: -0.002 min
Response: 14192839
Conc: 501.33 ng/ml




#19 AR-1242-4

R.T.: 5.679 min
Delta R.T.: -0.002 min
Response: 18731235
Conc: 518.43 ng/ml



#19 AR-1242-4


R.T.: 4.929 min
Delta R.T.: -0.002 min
Response: 15269722
Conc: 494.92 ng/ml



#20 AR-1242-5

R.T.: 6.419 min
Delta R.T.: -0.001 min
Response: 21241679
Conc: 507.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.452 min
Delta R.T.: -0.002 min
Response: 16563758
Conc: 497.27 ng/ml