

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022023\
 Data File : PP055804.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Feb 2023 15:59
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 21 00:54:53 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 20 05:24:03 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.326	3.569	83972030	72436437	52.116	49.608
2)	SA Decachlor...	10.080	8.577	107.3E6	57724529	52.586	50.881
Target Compounds							
16)	L4 AR-1242-1	5.498	4.651	25423509	18076802	526.516	491.310
17)	L4 AR-1242-2	5.520	4.670	36979421	26057987	520.405	494.724
18)	L4 AR-1242-3	5.582	4.847	23645772	14214108	530.005	502.079
19)	L4 AR-1242-4	5.681	4.930	19001307	15191909	525.908	492.399
20)	L4 AR-1242-5	6.421	5.454	21296591	16727081	508.403	502.174



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022023\
Data File : PP055804.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Feb 2023 15:59
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 21 00:54:53 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 20 05:24:03 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.326 min
Delta R.T.: 0.000 min
Response: 83972030
Conc: 52.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.569 min
Delta R.T.: 0.000 min
Response: 72436437
Conc: 49.61 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.080 min
Delta R.T.: 0.004 min
Response: 107273923
Conc: 52.59 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.577 min
Delta R.T.: 0.000 min
Response: 57724529
Conc: 50.88 ng/ml



#16 AR-1242-1


R.T.: 5.498 min
Delta R.T.: 0.000 min
Response: 25423509
Conc: 526.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.651 min
Delta R.T.: 0.000 min
Response: 18076802
Conc: 491.31 ng/ml




#17 AR-1242-2

R.T.: 5.520 min
Delta R.T.: 0.000 min
Response: 36979421
Conc: 520.40 ng/ml



#17 AR-1242-2


R.T.: 4.670 min
Delta R.T.: 0.000 min
Response: 26057987
Conc: 494.72 ng/ml



#18 AR-1242-3


R.T.: 5.582 min
Delta R.T.: 0.000 min
Response: 23645772
Conc: 530.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.847 min
Delta R.T.: 0.000 min
Response: 14214108
Conc: 502.08 ng/ml




#19 AR-1242-4

R.T.: 5.681 min
Delta R.T.: 0.000 min
Response: 19001307
Conc: 525.91 ng/ml



#19 AR-1242-4


R.T.: 4.930 min
Delta R.T.: 0.000 min
Response: 15191909
Conc: 492.40 ng/ml



#20 AR-1242-5

R.T.: 6.421 min
Delta R.T.: 0.000 min
Response: 21296591
Conc: 508.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.454 min
Delta R.T.: 0.000 min
Response: 16727081
Conc: 502.17 ng/ml