

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP022119.M
 Title : 8080.M
 Last Update : Fri Feb 22 02:24:04 2019
 Response Via : Initial Calibration

Calibration Files

750 =PP029170.D 1000 =PP029171.D 1500 =PP029172.D
 500 =PP029169.D 200 =PP029168.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	5.812	5.694	5.360	5.912	6.511	5.858	E4	7.17
2) T	3,5-DICHLOROBENZOI...	6.105	5.942	5.595	6.252	6.948	6.168	E4	8.10
3) T	4-Nitrophenol	6.800	6.652	6.320	6.954	7.527	6.850	E4	6.50
4) S	2,4-DCAA	3.062	3.132	2.869	2.986	2.979	3.006	E4	3.28
5) T	DICAMBA	1.313	1.288	1.232	1.333	1.483	1.330	E5	7.04
6) T	MCPP						0.000		-1.00
7) T	MCPA						0.000		-1.00
8) T	DICHLORPROP	3.139	3.073	2.948	3.111	3.156	3.086	E4	2.68
9) T	2,4-D	3.990	3.914	3.722	3.973	4.004	3.921	E4	2.96
10) T	Pentachlorophenol	6.501	6.452	6.356	6.413	6.860	6.516	E5	3.06
11) T	2,4,5-TP (SILVEX)	2.364	2.309	2.270	2.362	2.558	2.372	E5	4.67
12) T	2,4,5-T	2.113	2.064	2.021	2.093	2.265	2.111	E5	4.40
13) T	2,4-DB	1.917	1.843	1.811	1.859	2.121	1.910	E4	6.50
14) T	DINOSEB	1.472	1.429	1.410	1.458	1.512	1.456	E5	2.72
15) T	Picloram	2.829	2.772	2.682	2.749	2.871	2.781	E5	2.63
16) T	DCPA	2.827	2.742	2.710	2.826	3.019	2.825	E5	4.25

Signal #2 Calibration Files

750 =PP029170.D 1000 =PP029171.D 1500 =PP029172.D
 500 =PP029169.D 200 =PP029168.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	5.081	4.961	4.643	5.252	5.937	5.175	E4	9.29
2) T	3,5-DICHLOROBENZOI...	4.885	4.863	4.503	5.314	5.776	5.068	E4	9.65
3) T	4-Nitrophenol	2.441	2.292	2.129	2.505	2.607	2.395	E4	7.83
4) S	2,4-DCAA	3.663	3.661	3.406	3.982	4.264	3.795	E4	8.76
5) T	DICAMBA	1.231	1.237	1.167	1.338	1.428	1.280	E5	8.03
6) T	MCPP						0.000		-1.00
7) T	MCPA						0.000		-1.00
8) T	DICHLORPROP	3.266	3.236	3.031	3.539	3.833	3.381	E4	9.19
9) T	2,4-D	4.387	4.336	4.086	4.791	5.073	4.534	E4	8.67
10) T	Pentachlorophenol	4.967	5.020	4.810	5.400	5.537	5.147	E5	5.98
11) T	2,4,5-TP (SILVEX)	1.964	1.956	1.881	2.156	2.213	2.034	E5	7.00
12) T	2,4,5-T	1.641	1.609	1.554	1.703	1.771	1.656	E5	5.06
13) T	2,4-DB	2.245	2.195	2.141	2.346	2.336	2.253	E4	3.94
14) T	DINOSEB	1.305	1.288	1.256	1.349	1.356	1.311	E5	3.22
15) T	Picloram	2.308	2.267	2.200	2.206	2.271	2.250	E5	2.04
16) T	DCPA	2.302	2.251	2.218	2.384	2.465	2.324	E5	4.33

(#) = Out of Range