

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022121\
 Data File : PP033471.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Feb 2021 2:39
 Operator : DD\AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 22 05:15:16 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 22 05:14:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.720	3.989	7241984	4292742	95.828	96.640
2) SA Decachlor...	10.536	9.249	4733116	3185593	100.248	98.618
Target Compounds						
3) L1 AR-1016-1	6.019	5.234	1836842	1100674	950.272	935.915
4) L1 AR-1016-2	6.042	5.253	2812699	1706033	932.216	960.499
5) L1 AR-1016-3	6.107	5.445	1715157	920500	935.923	940.582
6) L1 AR-1016-4	6.212	5.495	1457810	702359	924.863	926.860
7) L1 AR-1016-5	6.525	5.723	1336891	905002	952.912	910.846
31) L7 AR-1260-1	7.694	6.812	2230448	1590166	941.916	936.387
32) L7 AR-1260-2	7.959	7.010	3259262	1836222	982.931	951.877
33) L7 AR-1260-3	8.323	7.163	2949562	1832044	1003.543	945.034
34) L7 AR-1260-4	8.551	7.643	2755369	1518841	1037.903	953.103
35) L7 AR-1260-5	8.864	7.892	6101286	3539291	1041.218	989.984



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022121\
Data File : PP033471.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Feb 2021 2:39
Operator : DD\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 22 05:15:16 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022121.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 22 05:14:45 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.720 min
Delta R.T.: -0.001 min
Response: 7241984
Conc: 95.83 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.989 min
Delta R.T.: 0.000 min
Response: 4292742
Conc: 96.64 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.536 min
Delta R.T.: 0.000 min
Response: 4733116
Conc: 100.25 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.249 min
Delta R.T.: 0.000 min
Response: 3185593
Conc: 98.62 ng/ml



#3 AR-1016-1


R.T.: 6.019 min
Delta R.T.: 0.000 min
Response: 1836842
Conc: 950.27 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.234 min
Delta R.T.: 0.000 min
Response: 1100674
Conc: 935.91 ng/ml




#4 AR-1016-2

R.T.: 6.042 min
Delta R.T.: 0.000 min
Response: 2812699
Conc: 932.22 ng/ml



#4 AR-1016-2


R.T.: 5.253 min
Delta R.T.: -0.001 min
Response: 1706033
Conc: 960.50 ng/ml



#5 AR-1016-3


R.T.: 6.107 min
Delta R.T.: 0.000 min
Response: 1715157
Conc: 935.92 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.445 min
Delta R.T.: 0.000 min
Response: 920500
Conc: 940.58 ng/ml




#6 AR-1016-4

R.T.: 6.212 min
Delta R.T.: 0.000 min
Response: 1457810
Conc: 924.86 ng/ml



#6 AR-1016-4


R.T.: 5.495 min
Delta R.T.: 0.000 min
Response: 702359
Conc: 926.86 ng/ml



#7 AR-1016-5


R.T.: 6.525 min
Delta R.T.: -0.001 min
Response: 1336891
Conc: 952.91 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.723 min
Delta R.T.: 0.000 min
Response: 905002
Conc: 910.85 ng/ml




#31 AR-1260-1

R.T.: 7.694 min
Delta R.T.: 0.000 min
Response: 2230448
Conc: 941.92 ng/ml



#31 AR-1260-1


R.T.: 6.812 min
Delta R.T.: 0.000 min
Response: 1590166
Conc: 936.39 ng/ml



#32 AR-1260-2


R.T.: 7.959 min
Delta R.T.: 0.000 min
Response: 3259262
Conc: 982.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#32 AR-1260-2

R.T.: 7.010 min
Delta R.T.: 0.000 min
Response: 1836222
Conc: 951.88 ng/ml




#33 AR-1260-3

R.T.: 8.323 min
Delta R.T.: 0.000 min
Response: 2949562
Conc: 1003.54 ng/ml



#33 AR-1260-3


R.T.: 7.163 min
Delta R.T.: 0.000 min
Response: 1832044
Conc: 945.03 ng/ml



#34 AR-1260-4


R.T.: 8.551 min
Delta R.T.: 0.000 min
Response: 2755369
Conc: 1037.90 ng/ml

Instrument :
ECD_P
ClientSampleId :




#34 AR-1260-4

R.T.: 7.643 min
Delta R.T.: 0.000 min
Response: 1518841
Conc: 953.10 ng/ml



#35 AR-1260-5

R.T.: 8.864 min
Delta R.T.: 0.000 min
Response: 6101286
Conc: 1041.22 ng/ml



#35 AR-1260-5

R.T.: 7.892 min
Delta R.T.: 0.000 min
Response: 3539291
Conc: 989.98 ng/ml