

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022122\
 Data File : PP043873.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Feb 2022 14:40
 Operator : YP\AJ
 Sample : N1601-24 10X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 22 00:13:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 11 04:48:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.913	3.171	2103884	2216108	0.952	1.180
2) SA Decachlor...	10.058	8.546	1505724	1593808	1.268	0.982
Target Compounds						
3) L1 AR-1016-1	5.177	4.325	183774	5620	2.533	0.092 #
4) L1 AR-1016-2	5.196	4.337	132339	55912	1.255	0.662 #
5) L1 AR-1016-3	5.251	4.524	225790	54444	3.429	1.165 #
6) L1 AR-1016-4	5.362	4.572	67441	467706	1.244	12.515 #
7) L1 AR-1016-5	5.683	4.795	184928	273881	3.521	5.995 #
8) L2 AR-1221-1	4.135	3.402	413763	54269	16.252	2.275 #
9) L2 AR-1221-2	4.250	3.487	304016	33825	15.836	1.873 #
10) L2 AR-1221-3	4.337f	3.577	271205	27870	4.721	0.540 #
11) L3 AR-1232-1	4.337f	3.577	271205	27870	5.588	0.648 #
12) L3 AR-1232-2	4.888	4.337	359005	55912	15.150	1.566 #
13) L3 AR-1232-3	5.196	4.524	132339	54444	3.074	2.784
14) L3 AR-1232-4	5.362	4.612	67441	257722	3.090	14.904 #
15) L3 AR-1232-5	5.463	4.795	563949	273881	34.964	14.580 #
16) L4 AR-1242-1	5.177	4.325	183774	5620	3.270	0.115 #
17) L4 AR-1242-2	5.196	4.337	132339	55912	1.600	0.830 #
18) L4 AR-1242-3	5.251	4.524	225790	54444	4.362	1.450 #
19) L4 AR-1242-4	5.362	4.612	67441	257722	1.579	7.137 #
20) L4 AR-1242-5	6.170	5.172	687378	1089646	15.918	25.312 #
21) L5 AR-1248-1	5.177	4.325	183774	5620	4.483	0.151 #
22) L5 AR-1248-2	5.463	4.572	563949	467706	9.926	9.560
23) L5 AR-1248-3	5.683	4.612	184928	257722	2.772	5.022 #
24) L5 AR-1248-4	0.000	4.795	0	273881	N.D.	4.672 #
25) L5 AR-1248-5	6.170	5.231	687378	20342	9.943	0.355 #
26) L6 AR-1254-1	6.095	5.172	1233829	1089646	16.330	12.477
27) L6 AR-1254-2	6.333	5.332	1512088	1400897	13.480	18.366 #
28) L6 AR-1254-3	6.734	5.764	1697942	1854987	14.674	15.136
29) L6 AR-1254-4	7.047	6.014	2020582	1166222	25.310	14.776 #
30) L6 AR-1254-5	7.508	6.465	1467513	1463718	17.111	12.657 #
31) L7 AR-1260-1	6.915	5.905	370645	1125782	4.426	14.078 #
32) L7 AR-1260-2	7.197	6.112	596135	1034875	6.586	10.597 #
33) L7 AR-1260-3	0.000	6.274	0	811890	N.D.	8.893 #
34) L7 AR-1260-4	7.822f	6.787	575570	197203	6.704	2.437 #
35) L7 AR-1260-5	8.181	7.055	527694	292874	3.404	1.584 #
36) L8 AR-1262-1	0.000	6.514	0	65343	N.D.	0.594 #
37) L8 AR-1262-2	8.181	6.787	527694	197203	3.269	1.933 #
38) L8 AR-1262-3	8.539	7.367	327146	48462	2.990	0.569 #
39) L8 AR-1262-4	8.649f	7.433	150792	310347	3.144	1.998 #
40) L8 AR-1262-5	0.000	7.986	0	52393	N.D.	0.673 #
41) L9 AR-1268-1	8.539f	7.367	327146	48462	1.678	0.191 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022122\
Data File : PP043873.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Feb 2022 14:40
Operator : YP\AJ
Sample : N1601-24 10X
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 22 00:13:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 11 04:48:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.649f	7.433	150792	310347	0.861	1.373 #
43)	L9 AR-1268-3	8.866	7.659	47034	23277	0.305	0.120 #
44)	L9 AR-1268-4	0.000	7.986	0	52393	N.D.	0.583 #
45)	L9 AR-1268-5	9.719	8.288	644169	80801	1.418	0.134 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022122\
Data File : PP043873.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Feb 2022 14:40
Operator : YP\AJ
Sample : N1601-24 10X
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 22 00:13:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 11 04:48:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.913 min
Delta R.T.: -0.006 min
Response: 2103884
Conc: 0.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.171 min
Delta R.T.: -0.003 min
Response: 2216108
Conc: 1.18 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.058 min
Delta R.T.: -0.014 min
Response: 1505724
Conc: 1.27 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.546 min
Delta R.T.: -0.008 min
Response: 1593808
Conc: 0.98 ng/ml



#3 AR-1016-1


R.T.: 5.177 min
Delta R.T.: 0.003 min
Response: 183774
Conc: 2.53 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.325 min
Delta R.T.: 0.002 min
Response: 5620
Conc: 0.09 ng/ml




#4 AR-1016-2

R.T.: 5.196 min
Delta R.T.: -0.002 min
Response: 132339
Conc: 1.25 ng/ml



#4 AR-1016-2


R.T.: 4.337 min
Delta R.T.: -0.004 min
Response: 55912
Conc: 0.66 ng/ml



#5 AR-1016-3


R.T.: 5.251 min
Delta R.T.: -0.012 min
Response: 225790
Conc: 3.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.524 min
Delta R.T.: -0.003 min
Response: 54444
Conc: 1.17 ng/ml




#6 AR-1016-4

R.T.: 5.362 min
Delta R.T.: -0.007 min
Response: 67441
Conc: 1.24 ng/ml



#6 AR-1016-4


R.T.: 4.572 min
Delta R.T.: -0.004 min
Response: 467706
Conc: 12.52 ng/ml



#7 AR-1016-5


R.T.: 5.683 min
Delta R.T.: -0.008 min
Response: 184928
Conc: 3.52 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 4.795 min
Delta R.T.: -0.006 min
Response: 273881
Conc: 6.00 ng/ml




#8 AR-1221-1

R.T.: 4.135 min
Delta R.T.: -0.012 min
Response: 413763
Conc: 16.25 ng/ml



#8 AR-1221-1


R.T.: 3.402 min
Delta R.T.: 0.000 min
Response: 54269
Conc: 2.28 ng/ml



#9 AR-1221-2


R.T.: 4.250 min
Delta R.T.: 0.012 min
Response: 304016
Conc: 15.84 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.487 min
Delta R.T.: -0.005 min
Response: 33825
Conc: 1.87 ng/ml




#10 AR-1221-3

R.T.: 4.337 min
Delta R.T.: 0.018 min
Response: 271205
Conc: 4.72 ng/ml



#10 AR-1221-3


R.T.: 3.577 min
Delta R.T.: 0.005 min
Response: 27870
Conc: 0.54 ng/ml



#11 AR-1232-1


R.T.: 4.337 min
Delta R.T.: 0.019 min
Response: 271205
Conc: 5.59 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.577 min
Delta R.T.: 0.005 min
Response: 27870
Conc: 0.65 ng/ml




#12 AR-1232-2

R.T.: 4.888 min
Delta R.T.: 0.004 min
Response: 359005
Conc: 15.15 ng/ml



#12 AR-1232-2


R.T.: 4.337 min
Delta R.T.: -0.005 min
Response: 55912
Conc: 1.57 ng/ml



#13 AR-1232-3


R.T.: 5.196 min
Delta R.T.: -0.002 min
Response: 132339
Conc: 3.07 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.524 min
Delta R.T.: -0.003 min
Response: 54444
Conc: 2.78 ng/ml




#14 AR-1232-4

R.T.: 5.362 min
Delta R.T.: -0.007 min
Response: 67441
Conc: 3.09 ng/ml



#14 AR-1232-4


R.T.: 4.612 min
Delta R.T.: -0.007 min
Response: 257722
Conc: 14.90 ng/ml



#15 AR-1232-5


R.T.: 5.463 min
Delta R.T.: -0.009 min
Response: 563949
Conc: 34.96 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 4.795 min
Delta R.T.: -0.006 min
Response: 273881
Conc: 14.58 ng/ml




#16 AR-1242-1

R.T.: 5.177 min
Delta R.T.: 0.002 min
Response: 183774
Conc: 3.27 ng/ml



#16 AR-1242-1


R.T.: 4.325 min
Delta R.T.: 0.002 min
Response: 5620
Conc: 0.12 ng/ml



#17 AR-1242-2


R.T.: 5.196 min
Delta R.T.: -0.001 min
Response: 132339
Conc: 1.60 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.337 min
Delta R.T.: -0.005 min
Response: 55912
Conc: 0.83 ng/ml




#18 AR-1242-3

R.T.: 5.251 min
Delta R.T.: -0.012 min
Response: 225790
Conc: 4.36 ng/ml



#18 AR-1242-3


R.T.: 4.524 min
Delta R.T.: -0.003 min
Response: 54444
Conc: 1.45 ng/ml



#19 AR-1242-4


R.T.: 5.362 min
Delta R.T.: -0.007 min
Response: 67441
Conc: 1.58 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.612 min
Delta R.T.: -0.008 min
Response: 257722
Conc: 7.14 ng/ml




#20 AR-1242-5

R.T.: 6.170 min
Delta R.T.: 0.000 min
Response: 687378
Conc: 15.92 ng/ml



#20 AR-1242-5


R.T.: 5.172 min
Delta R.T.: -0.008 min
Response: 1089646
Conc: 25.31 ng/ml



#21 AR-1248-1


R.T.: 5.177 min
Delta R.T.: 0.002 min
Response: 183774
Conc: 4.48 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.325 min
Delta R.T.: 0.003 min
Response: 5620
Conc: 0.15 ng/ml




#22 AR-1248-2

R.T.: 5.463 min
Delta R.T.: -0.009 min
Response: 563949
Conc: 9.93 ng/ml



#22 AR-1248-2


R.T.: 4.572 min
Delta R.T.: -0.004 min
Response: 467706
Conc: 9.56 ng/ml



#23 AR-1248-3


R.T.: 5.683 min
Delta R.T.: -0.007 min
Response: 184928
Conc: 2.77 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.612 min
Delta R.T.: -0.007 min
Response: 257722
Conc: 5.02 ng/ml




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 6.130 min
Response: 0
Conc: N.D.



#24 AR-1248-4


R.T.: 4.795 min
Delta R.T.: -0.006 min
Response: 273881
Conc: 4.67 ng/ml



#25 AR-1248-5


R.T.: 6.170 min
Delta R.T.: -0.001 min
Response: 687378
Conc: 9.94 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.231 min
Delta R.T.: 0.009 min
Response: 20342
Conc: 0.36 ng/ml




#26 AR-1254-1

R.T.: 6.095 min
Delta R.T.: -0.007 min
Response: 1233829
Conc: 16.33 ng/ml



#26 AR-1254-1


R.T.: 5.172 min
Delta R.T.: -0.008 min
Response: 1089646
Conc: 12.48 ng/ml



#27 AR-1254-2


R.T.: 6.333 min
Delta R.T.: -0.008 min
Response: 1512088
Conc: 13.48 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.332 min
Delta R.T.: -0.007 min
Response: 1400897
Conc: 18.37 ng/ml




#28 AR-1254-3

R.T.: 6.734 min
Delta R.T.: -0.007 min
Response: 1697942
Conc: 14.67 ng/ml



#28 AR-1254-3


R.T.: 5.764 min
Delta R.T.: -0.008 min
Response: 1854987
Conc: 15.14 ng/ml



#29 AR-1254-4


R.T.: 7.047 min
Delta R.T.: -0.008 min
Response: 2020582
Conc: 25.31 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.014 min
Delta R.T.: -0.008 min
Response: 1166222
Conc: 14.78 ng/ml




#30 AR-1254-5

R.T.: 7.508 min
Delta R.T.: -0.009 min
Response: 1467513
Conc: 17.11 ng/ml



#30 AR-1254-5


R.T.: 6.465 min
Delta R.T.: -0.009 min
Response: 1463718
Conc: 12.66 ng/ml



#31 AR-1260-1


R.T.: 6.915 min
Delta R.T.: -0.008 min
Response: 370645
Conc: 4.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 5.905 min
Delta R.T.: -0.008 min
Response: 1125782
Conc: 14.08 ng/ml




#32 AR-1260-2

R.T.: 7.197 min
Delta R.T.: -0.009 min
Response: 596135
Conc: 6.59 ng/ml



#32 AR-1260-2


R.T.: 6.112 min
Delta R.T.: -0.008 min
Response: 1034875
Conc: 10.60 ng/ml



#33 AR-1260-3


R.T.: 0.000 min
Exp R.T.: 7.601 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.274 min
Delta R.T.: -0.009 min
Response: 811890
Conc: 8.89 ng/ml




#34 AR-1260-4

R.T.: 7.822 min
Delta R.T.: -0.027 min
Response: 575570
Conc: 6.70 ng/ml



#34 AR-1260-4


R.T.: 6.787 min
Delta R.T.: -0.009 min
Response: 197203
Conc: 2.44 ng/ml



#35 AR-1260-5


R.T.: 8.181 min
Delta R.T.: -0.011 min
Response: 527694
Conc: 3.40 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.055 min
Delta R.T.: -0.008 min
Response: 292874
Conc: 1.58 ng/ml




#36 AR-1262-1

R.T.: 0.000 min
Exp R.T. : 7.602 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 6.514 min
Delta R.T.: -0.002 min
Response: 65343
Conc: 0.59 ng/ml



#37 AR-1262-2


R.T.: 8.181 min
Delta R.T.: -0.012 min
Response: 527694
Conc: 3.27 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.787 min
Delta R.T.: -0.010 min
Response: 197203
Conc: 1.93 ng/ml




#38 AR-1262-3

R.T.: 8.539 min
Delta R.T.: 0.013 min
Response: 327146
Conc: 2.99 ng/ml



#38 AR-1262-3


R.T.: 7.367 min
Delta R.T.: -0.005 min
Response: 48462
Conc: 0.57 ng/ml



#39 AR-1262-4


R.T.: 8.649 min
Delta R.T.: 0.030 min
Response: 150792
Conc: 3.14 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.433 min
Delta R.T.: -0.009 min
Response: 310347
Conc: 2.00 ng/ml




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T. : 9.303 min
Response: 0
Conc: N.D.



#40 AR-1262-5


R.T.: 7.986 min
Delta R.T.: 0.002 min
Response: 52393
Conc: 0.67 ng/ml



#41 AR-1268-1


R.T.: 8.539 min
Delta R.T.: 0.016 min
Response: 327146
Conc: 1.68 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.367 min
Delta R.T.: -0.005 min
Response: 48462
Conc: 0.19 ng/ml




#42 AR-1268-2

R.T.: 8.649 min
Delta R.T.: 0.028 min
Response: 150792
Conc: 0.86 ng/ml



#42 AR-1268-2


R.T.: 7.433 min
Delta R.T.: -0.010 min
Response: 310347
Conc: 1.37 ng/ml



#43 AR-1268-3


R.T.: 8.866 min
Delta R.T.: 0.007 min
Response: 47034
Conc: 0.30 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.659 min
Delta R.T.: -0.006 min
Response: 23277
Conc: 0.12 ng/ml




#44 AR-1268-4

R.T.: 0.000 min
Exp R.T. : 9.302 min
Response: 0
Conc: N.D.



#44 AR-1268-4


R.T.: 7.986 min
Delta R.T.: 0.002 min
Response: 52393
Conc: 0.58 ng/ml



#45 AR-1268-5

R.T.: 9.719 min
Delta R.T.: -0.009 min
Response: 644169
Conc: 1.42 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.288 min
Delta R.T.: -0.003 min
Response: 80801
Conc: 0.13 ng/ml