

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022122\
 Data File : PP043864.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Feb 2022 10:22
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 22 00:09:41 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 11 04:48:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.920	3.173	107.1E6	96497985	48.428	51.376
2) SA Decachlor...	10.065	8.548	62812258	76550636	52.877	47.184
Target Compounds						
16) L4 AR-1242-1	5.173	4.320	25576032	22730935	455.156	466.296
17) L4 AR-1242-2	5.197	4.339	37403311	31796825	452.091	472.219
18) L4 AR-1242-3	5.263	4.524	23485822	19080876	453.745	508.302
19) L4 AR-1242-4	5.368	4.616	19391109	17965934	453.992	497.540
20) L4 AR-1242-5	6.170	5.176	19665295	20945670	455.406	486.566



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022122\
Data File : PP043864.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Feb 2022 10:22
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 22 00:09:41 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 11 04:48:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.920 min
Delta R.T.: 0.001 min
Response: 107075927
Conc: 48.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.173 min
Delta R.T.: 0.000 min
Response: 96497985
Conc: 51.38 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.065 min
Delta R.T.: -0.006 min
Response: 62812258
Conc: 52.88 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.548 min
Delta R.T.: -0.007 min
Response: 76550636
Conc: 47.18 ng/ml



#16 AR-1242-1


R.T.: 5.173 min
Delta R.T.: 0.000 min
Response: 25576032
Conc: 455.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.320 min
Delta R.T.: -0.003 min
Response: 22730935
Conc: 466.30 ng/ml




#17 AR-1242-2

R.T.: 5.197 min
Delta R.T.: 0.000 min
Response: 37403311
Conc: 452.09 ng/ml



#17 AR-1242-2


R.T.: 4.339 min
Delta R.T.: -0.003 min
Response: 31796825
Conc: 472.22 ng/ml



#18 AR-1242-3


R.T.: 5.263 min
Delta R.T.: -0.001 min
Response: 23485822
Conc: 453.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.524 min
Delta R.T.: -0.003 min
Response: 19080876
Conc: 508.30 ng/ml




#19 AR-1242-4

R.T.: 5.368 min
Delta R.T.: -0.001 min
Response: 19391109
Conc: 453.99 ng/ml



#19 AR-1242-4


R.T.: 4.616 min
Delta R.T.: -0.004 min
Response: 17965934
Conc: 497.54 ng/ml



#20 AR-1242-5

R.T.: 6.170 min
Delta R.T.: 0.000 min
Response: 19665295
Conc: 455.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.176 min
Delta R.T.: -0.004 min
Response: 20945670
Conc: 486.57 ng/ml