

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022122\
 Data File : PP043866.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Feb 2022 10:53
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 22 00:10:28 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 11 04:48:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.916	3.173	106.3E6	97684231	48.068	52.008
2) SA Decachlor...	10.060	8.546	61603024	77599236	51.859	47.830
Target Compounds						
26) L6 AR-1254-1	6.095	5.174	34754056	41416454	459.974	474.238
27) L6 AR-1254-2	6.334	5.335	51573431	35774143	459.785	469.002
28) L6 AR-1254-3	6.734	5.766	53363177	57606046	461.170	470.030
29) L6 AR-1254-4	7.048	6.016	36988879	36765865	463.325	465.833
30) L6 AR-1254-5	7.509	6.468	40765348	54481530	475.325	471.112



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022122\
Data File : PP043866.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Feb 2022 10:53
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 22 00:10:28 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 11 04:48:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.916 min
Delta R.T.: -0.003 min
Response: 106279771
Conc: 48.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.173 min
Delta R.T.: -0.001 min
Response: 97684231
Conc: 52.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.060 min
Delta R.T.: -0.011 min
Response: 61603024
Conc: 51.86 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.546 min
Delta R.T.: -0.009 min
Response: 77599236
Conc: 47.83 ng/ml



#26 AR-1254-1


R.T.: 6.095 min
Delta R.T.: -0.007 min
Response: 34754056
Conc: 459.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.174 min
Delta R.T.: -0.005 min
Response: 41416454
Conc: 474.24 ng/ml




#27 AR-1254-2

R.T.: 6.334 min
Delta R.T.: -0.008 min
Response: 51573431
Conc: 459.79 ng/ml



#27 AR-1254-2


R.T.: 5.335 min
Delta R.T.: -0.005 min
Response: 35774143
Conc: 469.00 ng/ml



#28 AR-1254-3


R.T.: 6.734 min
Delta R.T.: -0.008 min
Response: 53363177
Conc: 461.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.766 min
Delta R.T.: -0.007 min
Response: 57606046
Conc: 470.03 ng/ml




#29 AR-1254-4

R.T.: 7.048 min
Delta R.T.: -0.008 min
Response: 36988879
Conc: 463.32 ng/ml



#29 AR-1254-4


R.T.: 6.016 min
Delta R.T.: -0.005 min
Response: 36765865
Conc: 465.83 ng/ml



#30 AR-1254-5

R.T.: 7.509 min
Delta R.T.: -0.008 min
Response: 40765348
Conc: 475.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.468 min
Delta R.T.: -0.007 min
Response: 54481530
Conc: 471.11 ng/ml