

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022222\
 Data File : PP043924.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Feb 2022 14:59
 Operator : YP\AJ
 Sample : N1626-11
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 CC-A6

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/23/2022
 Supervised By :Ankita Jodhani 02/23/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 23 00:34:49 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 11 04:48:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.916	3.172	35577208	34802010	16.091	18.529
2) SA Decachlor...	10.060	8.547	26789866	25789927	22.553	15.896 #
Target Compounds						
26) L6 AR-1254-1	6.098	5.174	45377185	53632864	600.573	614.122m
27) L6 AR-1254-2	6.336	5.332	66927194	37537834	596.666	492.124m
28) L6 AR-1254-3	6.735	5.765	62593178	60853767	540.937	496.529
29) L6 AR-1254-4	7.048	6.015	52585772	52011664	658.693	659.001
30) L6 AR-1254-5	7.510	6.467	74722940	96969243	871.272	838.511

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022222\
Data File : PP043924.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Feb 2022 14:59
Operator : YP\AJ
Sample : N1626-11
Misc :
ALS Vial : 18 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
CC-A6


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/23/2022
Supervised By :Ankita Jodhani 02/23/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 23 00:34:49 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 11 04:48:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.916 min
Delta R.T.: -0.003 min
Response: 35577208
Conc: 16.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
CC-A6


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/23/2022
Supervised By :Ankita Jodhani 02/23/2022




#1 Tetrachloro-m-xylene

R.T.: 3.172 min
Delta R.T.: -0.002 min
Response: 34802010
Conc: 18.53 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.060 min
Delta R.T.: 0.000 min
Response: 26789866
Conc: 22.55 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.547 min
Delta R.T.: -0.008 min
Response: 25789927
Conc: 15.90 ng/ml




#26 AR-1254-1

R.T.: 6.098 min
Delta R.T.: -0.004 min
Response: 45377185
Conc: 600.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
CC-A6


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/23/2022
Supervised By :Ankita Jodhani 02/23/2022




#26 AR-1254-1

R.T.: 5.174 min
Delta R.T.: -0.005 min
Response: 53632864
Conc: 614.12 ng/ml m




#27 AR-1254-2

R.T.: 6.336 min
Delta R.T.: -0.006 min
Response: 66927194
Conc: 596.67 ng/ml



#27 AR-1254-2

R.T.: 5.332 min
Delta R.T.: -0.007 min
Response: 37537834
Conc: 492.12 ng/ml m




#28 AR-1254-3

R.T.: 6.735 min
Delta R.T.: -0.007 min
Response: 62593178
Conc: 540.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
CC-A6


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/23/2022
Supervised By :Ankita Jodhani 02/23/2022




#28 AR-1254-3

R.T.: 5.765 min
Delta R.T.: -0.007 min
Response: 60853767
Conc: 496.53 ng/ml




#29 AR-1254-4

R.T.: 7.048 min
Delta R.T.: -0.007 min
Response: 52585772
Conc: 658.69 ng/ml



#29 AR-1254-4

R.T.: 6.015 min
Delta R.T.: -0.006 min
Response: 52011664
Conc: 659.00 ng/ml




#30 AR-1254-5

R.T.: 7.510 min
Delta R.T.: -0.007 min
Response: 74722940
Conc: 871.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
CC-A6

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/23/2022
Supervised By :Ankita Jodhani 02/23/2022



#30 AR-1254-5

R.T.: 6.467 min
Delta R.T.: -0.007 min
Response: 96969243
Conc: 838.51 ng/ml