

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022224\
 Data File : PP063341.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Feb 2024 22:02
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/23/2024
 Supervised By :Ankita Jodhani 02/23/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 23 00:17:36 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020124.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 05 12:13:11 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.491	3.643	108.4E6	113.4E6	45.336m	47.706
2)	SA Decachlor...	10.366	8.691	109.4E6	120.0E6	48.637m	48.425
Target Compounds							
3)	L1 AR-1016-1	5.674	4.737	35589698	35255915	472.947	492.814
4)	L1 AR-1016-2	5.697	4.756	51689103	50816745	475.909	494.743
5)	L1 AR-1016-3	5.760	4.933	34294754	27286513	496.959	493.436
6)	L1 AR-1016-4	5.859	4.975	27836013	22803231	497.042	460.790
7)	L1 AR-1016-5	6.157	5.191	28989038	29536992	474.760	469.545
31)	L7 AR-1260-1	7.293	6.230	57321389	65164479	470.973	491.617
32)	L7 AR-1260-2	7.553	6.420	62907753	74551206	494.642	504.740
33)	L7 AR-1260-3	7.915	6.574	52136899	70623912	504.360	487.134m
34)	L7 AR-1260-4	8.144	7.049	57694516	60502725	491.385	496.345
35)	L7 AR-1260-5	8.474	7.292	102.5E6	130.0E6	522.395	523.082

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022224\
Data File : PP063341.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Feb 2024 22:02
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/23/2024
Supervised By :Ankita Jodhani 02/23/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 23 00:17:36 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020124.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 05 12:13:11 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

