

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022322\
 Data File : PP043955.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Feb 2022 10:31
 Operator : YP\AJ
 Sample : N1630-01 10X
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 01

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/24/2022
 Supervised By :Ankita Jodhani 02/24/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 24 00:10:40 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 11 04:48:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.915	3.170	2289018	1197072	1.035m	0.637m#
2) SA Decachlor...	10.051	8.544	2405815	870904	2.025m	0.537m#
Target Compounds						
26) L6 AR-1254-1	6.093	5.171	34431033	44753180	455.699	512.445
27) L6 AR-1254-2	6.334	5.332	88047293	61475554	784.955	805.949
28) L6 AR-1254-3	6.734	5.764	130.3E6	127.2E6	1125.691	1037.532
29) L6 AR-1254-4	7.047	6.014	129.3E6	113.9E6	1619.697	1443.463
30) L6 AR-1254-5	7.508	6.466	217.7E6	273.8E6	2538.919	2368.029

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022322\
Data File : PP043955.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Feb 2022 10:31
Operator : YP\AJ
Sample : N1630-01 10X
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
01

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/24/2022
Supervised By :Ankita Jodhani 02/24/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 24 00:10:40 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 11 04:48:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

