

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022323\
 Data File : PP055875.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Feb 2023 19:44
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 23 23:20:55 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 20 05:24:03 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.324	3.567	85049877	66983654	52.785	45.874
2) SA Decachlor...	10.073	8.571	89754933	55933285	43.998	49.302
Target Compounds						
3) L1 AR-1016-1	5.495	4.649	30612922	20865077	513.005	463.359
4) L1 AR-1016-2	5.517	4.667	43904438	29708933	501.827	457.455
5) L1 AR-1016-3	5.579	4.844	27905755	16095261	502.356	463.133
6) L1 AR-1016-4	5.678	4.886	22535652	13390499	505.721	447.981
7) L1 AR-1016-5	5.974	5.099	23672573	18596265	490.271	470.392
31) L7 AR-1260-1	7.105	6.134	51505859	34048434	497.430	467.047
32) L7 AR-1260-2	7.364	6.323	59614237	39074045	473.987	478.861
33) L7 AR-1260-3	7.725	6.477	43749951	36833404	498.137	435.865
34) L7 AR-1260-4	7.951	6.949	49206643	30815304	457.843	481.484
35) L7 AR-1260-5	8.268	7.193	95329157	65593446	461.756	484.587



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022323\
Data File : PP055875.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Feb 2023 19:44
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 23 23:20:55 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 20 05:24:03 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.324 min
Delta R.T.: 0.000 min
Response: 85049877
Conc: 52.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.567 min
Delta R.T.: -0.001 min
Response: 66983654
Conc: 45.87 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.073 min
Delta R.T.: -0.002 min
Response: 89754933
Conc: 44.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.571 min
Delta R.T.: -0.005 min
Response: 55933285
Conc: 49.30 ng/ml



#3 AR-1016-1


R.T.: 5.495 min
Delta R.T.: -0.001 min
Response: 30612922
Conc: 513.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.649 min
Delta R.T.: -0.003 min
Response: 20865077
Conc: 463.36 ng/ml




#4 AR-1016-2

R.T.: 5.517 min
Delta R.T.: -0.001 min
Response: 43904438
Conc: 501.83 ng/ml



#4 AR-1016-2


R.T.: 4.667 min
Delta R.T.: -0.003 min
Response: 29708933
Conc: 457.46 ng/ml



#5 AR-1016-3


R.T.: 5.579 min
Delta R.T.: -0.002 min
Response: 27905755
Conc: 502.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.844 min
Delta R.T.: -0.003 min
Response: 16095261
Conc: 463.13 ng/ml




#6 AR-1016-4

R.T.: 5.678 min
Delta R.T.: -0.002 min
Response: 22535652
Conc: 505.72 ng/ml



#6 AR-1016-4


R.T.: 4.886 min
Delta R.T.: -0.003 min
Response: 13390499
Conc: 447.98 ng/ml



#7 AR-1016-5


R.T.: 5.974 min
Delta R.T.: -0.002 min
Response: 23672573
Conc: 490.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.099 min
Delta R.T.: -0.003 min
Response: 18596265
Conc: 470.39 ng/ml




#31 AR-1260-1

R.T.: 7.105 min
Delta R.T.: -0.002 min
Response: 51505859
Conc: 497.43 ng/ml



#31 AR-1260-1


R.T.: 6.134 min
Delta R.T.: -0.004 min
Response: 34048434
Conc: 467.05 ng/ml



#32 AR-1260-2


R.T.: 7.364 min
Delta R.T.: 0.000 min
Response: 59614237
Conc: 473.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.323 min
Delta R.T.: -0.004 min
Response: 39074045
Conc: 478.86 ng/ml




#33 AR-1260-3

R.T.: 7.725 min
Delta R.T.: -0.002 min
Response: 43749951
Conc: 498.14 ng/ml



#33 AR-1260-3


R.T.: 6.477 min
Delta R.T.: -0.003 min
Response: 36833404
Conc: 435.87 ng/ml



#34 AR-1260-4


R.T.: 7.951 min
Delta R.T.: -0.002 min
Response: 49206643
Conc: 457.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.949 min
Delta R.T.: -0.005 min
Response: 30815304
Conc: 481.48 ng/ml



#35 AR-1260-5

R.T.: 8.268 min
Delta R.T.: 0.000 min
Response: 95329157
Conc: 461.76 ng/ml



#35 AR-1260-5

R.T.: 7.193 min
Delta R.T.: -0.004 min
Response: 65593446
Conc: 484.59 ng/ml