

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022324\
 Data File : PP063353.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Feb 2024 12:14
 Operator : YP\AJ
 Sample : P1575-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PTP-01-(0.5-1.0)

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/26/2024
 Supervised By :Ankita Jodhani 02/26/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 23 13:29:52 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020124.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 05 12:13:11 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.492	3.643	31976302	30006452	13.368	12.624
2)	SA Decachlor...	10.369	8.692	39816507	38505139	17.705	15.544
Target Compounds							
26)	L6 AR-1254-1	6.539	5.546	8900114	10009578	99.736m	82.760
27)	L6 AR-1254-2	6.761	5.694	12083174	8341666	91.626	77.172
28)	L6 AR-1254-3	7.130	6.100	16301220	20928857	118.582	123.749
29)	L6 AR-1254-4	7.418	6.329	8972241	7174356	109.191	89.900
30)	L6 AR-1254-5	7.840	6.751	24566714	24985207	238.640	177.348 #
31)	L7 AR-1260-1	7.296	6.232	13469989	14718888	110.674	111.043
32)	L7 AR-1260-2	7.554	6.420	18043413	16074773	141.875	108.832
33)	L7 AR-1260-3	7.917	6.574	7810830	13967235	75.560	96.340m#
34)	L7 AR-1260-4	8.138	7.050	14470467	6670566	123.245m	54.723 #
35)	L7 AR-1260-5	8.477	7.292	15381897	19964281	78.379	80.355

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022324\
Data File : PP063353.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Feb 2024 12:14
Operator : YP\AJ
Sample : P1575-01
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PTP-01-(0.5-1.0)

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/26/2024
Supervised By :Ankita Jodhani 02/26/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 23 13:29:52 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020124.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 05 12:13:11 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

