

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022324\
 Data File : PP063388.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Feb 2024 23:05
 Operator : YP\AJ
 Sample : P1574-07
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SB-7

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/26/2024
 Supervised By :Ankita Jodhani 02/26/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 24 00:26:34 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020124.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 05 12:13:11 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.491	3.642	47586507	51468117	19.894	21.654
2)	SA Decachlor...	10.369	8.691	61499060	60009301	27.347	24.226
Target Compounds							
21)	L5 AR-1248-1	5.673	4.737	26474899	26012670	596.953m	627.893
22)	L5 AR-1248-2	5.951	4.975	35693168	33257170	496.818m	502.675m
23)	L5 AR-1248-3	6.157	5.017	60487751	38962667	778.405m	556.185m#
24)	L5 AR-1248-4	6.565	5.189	37323759	51552083	500.832m	645.133m#
25)	L5 AR-1248-5	6.604	5.585	51648744	33000745	663.929m	480.980 #
31)	L7 AR-1260-1	7.294	6.230	7583866	12856313	62.312m	96.991 #
32)	L7 AR-1260-2	7.554	6.419	5451158	6384025	42.862m	43.222
33)	L7 AR-1260-3	7.916	6.572	2949356	5253877	28.531	36.239m#
34)	L7 AR-1260-4	8.143	7.049	3263938	2036361	27.799m	16.706 #
35)	L7 AR-1260-5	8.476	7.291	2987562	5016638	15.223	20.192 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022324\
Data File : PP063388.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Feb 2024 23:05
Operator : YP\AJ
Sample : P1574-07
Misc :
ALS Vial : 40 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SB-7

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/26/2024
Supervised By :Ankita Jodhani 02/26/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 24 00:26:34 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP020124.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 05 12:13:11 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

