

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022421\
 Data File : PP033551.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Feb 2021 18:54
 Operator : DD\AJ
 Sample : M1481-04
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TT-027-IDWGW-20210223

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 25 05:09:19 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 22 12:13:48 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.721	3.993	1549864	868791	21.057	19.728
2) SA Decachlor...	10.535	9.250	261584	181548	5.624	5.746

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022421\
Data File : PP033551.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Feb 2021 18:54
Operator : DD\AJ
Sample : M1481-04
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TT-027-IDWGW-20210223

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 25 05:09:19 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022121.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 22 12:13:48 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.721 min
Delta R.T.: 0.001 min
Response: 1549864
Conc: 21.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
TT-027-IDW/GW-20210223




#1 Tetrachloro-m-xylene

R.T.: 3.993 min
Delta R.T.: 0.002 min
Response: 868791
Conc: 19.73 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.535 min
Delta R.T.: 0.000 min
Response: 261584
Conc: 5.62 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.250 min
Delta R.T.: 0.000 min
Response: 181548
Conc: 5.75 ng/ml