

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022422\
 Data File : PP043975.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Feb 2022 19:12
 Operator : YP\AJ
 Sample : PB142891BL
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB142891BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 25 00:39:28 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 11 04:48:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.916	3.173	40198454	38644354	18.181	20.574
2)	SA Decachlor...	10.056	8.542	24418726	31524299	20.556	19.431
Target Compounds							
3)	L1 AR-1016-1	5.170	4.325	22559	6264	0.311	0.102 #
4)	L1 AR-1016-2	5.204	4.340	14675	54321	0.139	0.643 #
5)	L1 AR-1016-3	5.263	4.529	44229	19572	0.672	0.419 #
6)	L1 AR-1016-4	5.379	4.575	355237	28453	6.551	0.761 #
7)	L1 AR-1016-5	5.688	4.802	140906	40432	2.683	0.885 #
8)	L2 AR-1221-1	4.136	3.408	668150	8505	26.244	0.357 #
9)	L2 AR-1221-2	4.240	3.485	914724	798085	47.646	44.191
10)	L2 AR-1221-3	4.324	3.567	138986	70032	2.419	1.358 #
11)	L3 AR-1232-1	4.324	3.567	138986	70032	2.864	1.628 #
12)	L3 AR-1232-2	4.891	4.340	95500	54321	4.030	1.522 #
13)	L3 AR-1232-3	5.190	4.529	66337	19572	1.541	1.001 #
14)	L3 AR-1232-4	5.379	4.621	355237	70537	16.274	4.079 #
15)	L3 AR-1232-5	5.471	4.802	219566	40432	13.613	2.152 #
16)	L4 AR-1242-1	5.170	4.325	22559	6264	0.401	0.128 #
17)	L4 AR-1242-2	5.190	4.340	66337	54321	0.802	0.807
18)	L4 AR-1242-3	5.263	4.529	44229	19572	0.855	0.521 #
19)	L4 AR-1242-4	5.379	4.621	355237	70537	8.317	1.953 #
20)	L4 AR-1242-5	6.162	5.168	188181	181555	4.358	4.218
21)	L5 AR-1248-1	5.170	4.320	22559	24807	0.550	0.666
22)	L5 AR-1248-2	5.471	4.575	219566	28453	3.865	0.582 #
23)	L5 AR-1248-3	5.688	4.621	140906	70537	2.112	1.375 #
24)	L5 AR-1248-4	6.124	4.802	59922	40432	0.858	0.690
25)	L5 AR-1248-5	6.162	5.222	188181	110604	2.722	1.930 #
26)	L6 AR-1254-1	6.093	5.168	247940	181555	3.282	2.079 #
27)	L6 AR-1254-2	6.343	5.335	224364	203551	2.000	2.669 #
28)	L6 AR-1254-3	6.739	5.777	76164	40602	0.658	0.331 #
29)	L6 AR-1254-4	7.056	6.016	838301	10374	10.501	0.131 #
30)	L6 AR-1254-5	0.000	6.479	0	9666	N.D.	0.084 #
31)	L7 AR-1260-1	6.914	5.913	37107	41600	0.443	0.520
32)	L7 AR-1260-2	7.202	6.115	1773516	12196	19.592	0.125 #
33)	L7 AR-1260-3	7.608	6.284	633800	22799	8.743	0.250 #
34)	L7 AR-1260-4	7.867f	6.805	541753	90370	6.310	1.117 #
35)	L7 AR-1260-5	8.202	7.069	47786	15120	0.308	0.082 #
36)	L8 AR-1262-1	7.608	6.518	633800	47364	6.419	0.431 #
37)	L8 AR-1262-2	8.202	6.805	47786	90370	0.296	0.886 #
38)	L8 AR-1262-3	8.530	7.367	33157	22469	0.303	0.264
39)	L8 AR-1262-4	8.616	7.445	29643	19341	0.618	0.125 #
41)	L9 AR-1268-1	8.530	7.367	33157	22469	0.170	0.089 #
42)	L9 AR-1268-2	8.616	7.445	29643	19341	0.169	0.086 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022422\
 Data File : PP043975.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Feb 2022 19:12
 Operator : YP\AJ
 Sample : PB142891BL
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB142891BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 25 00:39:28 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 11 04:48:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
43) L9	AR-1268-3	8.851	7.668	58319	5826	0.378	0.030 #
45) L9	AR-1268-5	9.714	8.298	144097	27342	0.317	0.046 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022422\
Data File : PP043975.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Feb 2022 19:12
Operator : YP\AJ
Sample : PB142891BL
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB142891BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 25 00:39:28 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 11 04:48:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.916 min
Delta R.T.: -0.003 min
Response: 40198454
Conc: 18.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142891BL




#1 Tetrachloro-m-xylene

R.T.: 3.173 min
Delta R.T.: -0.001 min
Response: 38644354
Conc: 20.57 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.056 min
Delta R.T.: -0.004 min
Response: 24418726
Conc: 20.56 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.542 min
Delta R.T.: -0.012 min
Response: 31524299
Conc: 19.43 ng/ml



#3 AR-1016-1


R.T.: 5.170 min
Delta R.T.: -0.004 min
Response: 22559
Conc: 0.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142891BL




#3 AR-1016-1

R.T.: 4.325 min
Delta R.T.: 0.002 min
Response: 6264
Conc: 0.10 ng/ml




#4 AR-1016-2

R.T.: 5.204 min
Delta R.T.: 0.007 min
Response: 14675
Conc: 0.14 ng/ml



#4 AR-1016-2


R.T.: 4.340 min
Delta R.T.: -0.001 min
Response: 54321
Conc: 0.64 ng/ml



#5 AR-1016-3


R.T.: 5.263 min
Delta R.T.: 0.000 min
Response: 44229
Conc: 0.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142891BL




#5 AR-1016-3

R.T.: 4.529 min
Delta R.T.: 0.001 min
Response: 19572
Conc: 0.42 ng/ml




#6 AR-1016-4

R.T.: 5.379 min
Delta R.T.: 0.009 min
Response: 355237
Conc: 6.55 ng/ml



#6 AR-1016-4


R.T.: 4.575 min
Delta R.T.: 0.000 min
Response: 28453
Conc: 0.76 ng/ml



#7 AR-1016-5


R.T.: 5.688 min
Delta R.T.: -0.003 min
Response: 140906
Conc: 2.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142891BL




#7 AR-1016-5

R.T.: 4.802 min
Delta R.T.: 0.000 min
Response: 40432
Conc: 0.89 ng/ml




#8 AR-1221-1

R.T.: 4.136 min
Delta R.T.: -0.011 min
Response: 668150
Conc: 26.24 ng/ml



#8 AR-1221-1


R.T.: 3.408 min
Delta R.T.: 0.005 min
Response: 8505
Conc: 0.36 ng/ml



#9 AR-1221-2


R.T.: 4.240 min
Delta R.T.: 0.003 min
Response: 914724
Conc: 47.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142891BL




#9 AR-1221-2

R.T.: 3.485 min
Delta R.T.: -0.008 min
Response: 798085
Conc: 44.19 ng/ml




#10 AR-1221-3

R.T.: 4.324 min
Delta R.T.: 0.006 min
Response: 138986
Conc: 2.42 ng/ml



#10 AR-1221-3


R.T.: 3.567 min
Delta R.T.: -0.005 min
Response: 70032
Conc: 1.36 ng/ml



#11 AR-1232-1


R.T.: 4.324 min
Delta R.T.: 0.006 min
Response: 138986
Conc: 2.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142891BL




#11 AR-1232-1

R.T.: 3.567 min
Delta R.T.: -0.005 min
Response: 70032
Conc: 1.63 ng/ml




#12 AR-1232-2

R.T.: 4.891 min
Delta R.T.: 0.007 min
Response: 95500
Conc: 4.03 ng/ml



#12 AR-1232-2


R.T.: 4.340 min
Delta R.T.: -0.001 min
Response: 54321
Conc: 1.52 ng/ml



#13 AR-1232-3


R.T.: 5.190 min
Delta R.T.: -0.007 min
Response: 66337
Conc: 1.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142891BL




#13 AR-1232-3

R.T.: 4.529 min
Delta R.T.: 0.002 min
Response: 19572
Conc: 1.00 ng/ml




#14 AR-1232-4

R.T.: 5.379 min
Delta R.T.: 0.009 min
Response: 355237
Conc: 16.27 ng/ml



#14 AR-1232-4


R.T.: 4.621 min
Delta R.T.: 0.002 min
Response: 70537
Conc: 4.08 ng/ml



#15 AR-1232-5


R.T.: 5.471 min
Delta R.T.: 0.000 min
Response: 219566
Conc: 13.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142891BL




#15 AR-1232-5

R.T.: 4.802 min
Delta R.T.: 0.001 min
Response: 40432
Conc: 2.15 ng/ml




#16 AR-1242-1

R.T.: 5.170 min
Delta R.T.: -0.004 min
Response: 22559
Conc: 0.40 ng/ml



#16 AR-1242-1


R.T.: 4.325 min
Delta R.T.: 0.002 min
Response: 6264
Conc: 0.13 ng/ml



#17 AR-1242-2


R.T.: 5.190 min
Delta R.T.: -0.007 min
Response: 66337
Conc: 0.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142891BL




#17 AR-1242-2

R.T.: 4.340 min
Delta R.T.: -0.002 min
Response: 54321
Conc: 0.81 ng/ml




#18 AR-1242-3

R.T.: 5.263 min
Delta R.T.: 0.000 min
Response: 44229
Conc: 0.85 ng/ml



#18 AR-1242-3


R.T.: 4.529 min
Delta R.T.: 0.002 min
Response: 19572
Conc: 0.52 ng/ml



#19 AR-1242-4


R.T.: 5.379 min
Delta R.T.: 0.009 min
Response: 355237
Conc: 8.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142891BL




#19 AR-1242-4

R.T.: 4.621 min
Delta R.T.: 0.001 min
Response: 70537
Conc: 1.95 ng/ml




#20 AR-1242-5

R.T.: 6.162 min
Delta R.T.: -0.009 min
Response: 188181
Conc: 4.36 ng/ml



#20 AR-1242-5


R.T.: 5.168 min
Delta R.T.: -0.012 min
Response: 181555
Conc: 4.22 ng/ml



#21 AR-1248-1


R.T.: 5.170 min
Delta R.T.: -0.004 min
Response: 22559
Conc: 0.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142891BL




#21 AR-1248-1

R.T.: 4.320 min
Delta R.T.: -0.003 min
Response: 24807
Conc: 0.67 ng/ml




#22 AR-1248-2

R.T.: 5.471 min
Delta R.T.: 0.000 min
Response: 219566
Conc: 3.86 ng/ml



#22 AR-1248-2


R.T.: 4.575 min
Delta R.T.: 0.000 min
Response: 28453
Conc: 0.58 ng/ml



#23 AR-1248-3


R.T.: 5.688 min
Delta R.T.: -0.002 min
Response: 140906
Conc: 2.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142891BL




#23 AR-1248-3

R.T.: 4.621 min
Delta R.T.: 0.002 min
Response: 70537
Conc: 1.37 ng/ml




#24 AR-1248-4

R.T.: 6.124 min
Delta R.T.: -0.007 min
Response: 59922
Conc: 0.86 ng/ml



#24 AR-1248-4


R.T.: 4.802 min
Delta R.T.: 0.001 min
Response: 40432
Conc: 0.69 ng/ml



#25 AR-1248-5


R.T.: 6.162 min
Delta R.T.: -0.009 min
Response: 188181
Conc: 2.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142891BL




#25 AR-1248-5

R.T.: 5.222 min
Delta R.T.: 0.000 min
Response: 110604
Conc: 1.93 ng/ml




#26 AR-1254-1

R.T.: 6.093 min
Delta R.T.: -0.009 min
Response: 247940
Conc: 3.28 ng/ml



#26 AR-1254-1


R.T.: 5.168 min
Delta R.T.: -0.012 min
Response: 181555
Conc: 2.08 ng/ml



#27 AR-1254-2


R.T.: 6.343 min
Delta R.T.: 0.001 min
Response: 224364
Conc: 2.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142891BL




#27 AR-1254-2

R.T.: 5.335 min
Delta R.T.: -0.004 min
Response: 203551
Conc: 2.67 ng/ml




#28 AR-1254-3

R.T.: 6.739 min
Delta R.T.: -0.002 min
Response: 76164
Conc: 0.66 ng/ml



#28 AR-1254-3


R.T.: 5.777 min
Delta R.T.: 0.004 min
Response: 40602
Conc: 0.33 ng/ml



#29 AR-1254-4


R.T.: 7.056 min
Delta R.T.: 0.000 min
Response: 838301
Conc: 10.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142891BL




#29 AR-1254-4

R.T.: 6.016 min
Delta R.T.: -0.006 min
Response: 10374
Conc: 0.13 ng/ml




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 7.517 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 6.479 min
Delta R.T.: 0.005 min
Response: 9666
Conc: 0.08 ng/ml



#31 AR-1260-1


R.T.: 6.914 min
Delta R.T.: -0.009 min
Response: 37107
Conc: 0.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142891BL




#31 AR-1260-1

R.T.: 5.913 min
Delta R.T.: 0.000 min
Response: 41600
Conc: 0.52 ng/ml




#32 AR-1260-2

R.T.: 7.202 min
Delta R.T.: -0.004 min
Response: 1773516
Conc: 19.59 ng/ml



#32 AR-1260-2


R.T.: 6.115 min
Delta R.T.: -0.004 min
Response: 12196
Conc: 0.12 ng/ml



#33 AR-1260-3


R.T.: 7.608 min
Delta R.T.: 0.007 min
Response: 633800
Conc: 8.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142891BL




#33 AR-1260-3

R.T.: 6.284 min
Delta R.T.: 0.001 min
Response: 22799
Conc: 0.25 ng/ml




#34 AR-1260-4

R.T.: 7.867 min
Delta R.T.: 0.018 min
Response: 541753
Conc: 6.31 ng/ml



#34 AR-1260-4


R.T.: 6.805 min
Delta R.T.: 0.009 min
Response: 90370
Conc: 1.12 ng/ml



#35 AR-1260-5


R.T.: 8.202 min
Delta R.T.: 0.010 min
Response: 47786
Conc: 0.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142891BL




#35 AR-1260-5

R.T.: 7.069 min
Delta R.T.: 0.005 min
Response: 15120
Conc: 0.08 ng/ml




#36 AR-1262-1

R.T.: 7.608 min
Delta R.T.: 0.006 min
Response: 633800
Conc: 6.42 ng/ml



#36 AR-1262-1


R.T.: 6.518 min
Delta R.T.: 0.002 min
Response: 47364
Conc: 0.43 ng/ml



#37 AR-1262-2


R.T.: 8.202 min
Delta R.T.: 0.009 min
Response: 47786
Conc: 0.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142891BL




#37 AR-1262-2

R.T.: 6.805 min
Delta R.T.: 0.008 min
Response: 90370
Conc: 0.89 ng/ml




#38 AR-1262-3

R.T.: 8.530 min
Delta R.T.: 0.003 min
Response: 33157
Conc: 0.30 ng/ml



#38 AR-1262-3


R.T.: 7.367 min
Delta R.T.: -0.005 min
Response: 22469
Conc: 0.26 ng/ml



#39 AR-1262-4


R.T.: 8.616 min
Delta R.T.: -0.003 min
Response: 29643
Conc: 0.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142891BL




#39 AR-1262-4

R.T.: 7.445 min
Delta R.T.: 0.004 min
Response: 19341
Conc: 0.12 ng/ml




#40 AR-1262-5

R.T.: 9.305 min
Delta R.T.: 0.002 min
Response: -57745
Conc: N.D.



#40 AR-1262-5


R.T.: 7.987 min
Delta R.T.: 0.003 min
Response: 63458
Conc: 0.82 ng/ml



#41 AR-1268-1


R.T.: 8.530 min
Delta R.T.: 0.007 min
Response: 33157
Conc: 0.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142891BL




#41 AR-1268-1

R.T.: 7.367 min
Delta R.T.: -0.005 min
Response: 22469
Conc: 0.09 ng/ml




#42 AR-1268-2

R.T.: 8.616 min
Delta R.T.: -0.006 min
Response: 29643
Conc: 0.17 ng/ml



#42 AR-1268-2


R.T.: 7.445 min
Delta R.T.: 0.003 min
Response: 19341
Conc: 0.09 ng/ml



#43 AR-1268-3


R.T.: 8.851 min
Delta R.T.: -0.008 min
Response: 58319
Conc: 0.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142891BL




#43 AR-1268-3

R.T.: 7.668 min
Delta R.T.: 0.003 min
Response: 5826
Conc: 0.03 ng/ml




#44 AR-1268-4

R.T.: 9.305 min
Delta R.T.: 0.002 min
Response: -57745
Conc: N.D.



#44 AR-1268-4


R.T.: 7.987 min
Delta R.T.: 0.003 min
Response: 63458
Conc: 0.71 ng/ml



#45 AR-1268-5

R.T.: 9.714 min
Delta R.T.: -0.014 min
Response: 144097
Conc: 0.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB142891BL



#45 AR-1268-5

R.T.: 8.298 min
Delta R.T.: 0.007 min
Response: 27342
Conc: 0.05 ng/ml