

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022423\
 Data File : PP055968.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Feb 2023 01:55
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 25 07:54:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 20 05:24:03 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.322	3.565	90836675	70533938	56.377	48.305
2) SA Decachlor...	10.062	8.565	100.1E6	55747082	49.078	49.138
Target Compounds						
3) L1 AR-1016-1	5.491	4.646	32931897	21319465	551.865	473.450
4) L1 AR-1016-2	5.514	4.664	47539542	30537620	543.376	470.215
5) L1 AR-1016-3	5.576	4.840	30188284	16361700	543.446	470.799
6) L1 AR-1016-4	5.675	4.882	24431384	13545976	548.263	453.183
7) L1 AR-1016-5	5.971	5.095	25640578	18172066	531.030	459.662
31) L7 AR-1260-1	7.101	6.130	55304193	34044382	534.113	466.991
32) L7 AR-1260-2	7.359	6.319	62711170	39154085	498.611	479.842
33) L7 AR-1260-3	7.720	6.471	46973355	36983778	534.838	437.645
34) L7 AR-1260-4	7.946	6.944	54423957	31410435	506.387	490.782
35) L7 AR-1260-5	8.262	7.188	106.8E6	66790401	517.249	493.430



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022423\
Data File : PP055968.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Feb 2023 01:55
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 25 07:54:45 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 20 05:24:03 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.322 min
Delta R.T.: -0.003 min
Response: 90836675
Conc: 56.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.565 min
Delta R.T.: -0.004 min
Response: 70533938
Conc: 48.31 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.062 min
Delta R.T.: -0.014 min
Response: 100116173
Conc: 49.08 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.565 min
Delta R.T.: -0.012 min
Response: 55747082
Conc: 49.14 ng/ml



#3 AR-1016-1


R.T.: 5.491 min
Delta R.T.: -0.005 min
Response: 32931897
Conc: 551.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.646 min
Delta R.T.: -0.006 min
Response: 21319465
Conc: 473.45 ng/ml




#4 AR-1016-2

R.T.: 5.514 min
Delta R.T.: -0.005 min
Response: 47539542
Conc: 543.38 ng/ml



#4 AR-1016-2


R.T.: 4.664 min
Delta R.T.: -0.006 min
Response: 30537620
Conc: 470.22 ng/ml



#5 AR-1016-3


R.T.: 5.576 min
Delta R.T.: -0.005 min
Response: 30188284
Conc: 543.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.840 min
Delta R.T.: -0.007 min
Response: 16361700
Conc: 470.80 ng/ml




#6 AR-1016-4

R.T.: 5.675 min
Delta R.T.: -0.005 min
Response: 24431384
Conc: 548.26 ng/ml



#6 AR-1016-4


R.T.: 4.882 min
Delta R.T.: -0.007 min
Response: 13545976
Conc: 453.18 ng/ml



#7 AR-1016-5


R.T.: 5.971 min
Delta R.T.: -0.005 min
Response: 25640578
Conc: 531.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.095 min
Delta R.T.: -0.007 min
Response: 18172066
Conc: 459.66 ng/ml




#31 AR-1260-1

R.T.: 7.101 min
Delta R.T.: -0.007 min
Response: 55304193
Conc: 534.11 ng/ml



#31 AR-1260-1


R.T.: 6.130 min
Delta R.T.: -0.008 min
Response: 34044382
Conc: 466.99 ng/ml



#32 AR-1260-2


R.T.: 7.359 min
Delta R.T.: -0.006 min
Response: 62711170
Conc: 498.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.319 min
Delta R.T.: -0.008 min
Response: 39154085
Conc: 479.84 ng/ml




#33 AR-1260-3

R.T.: 7.720 min
Delta R.T.: -0.007 min
Response: 46973355
Conc: 534.84 ng/ml



#33 AR-1260-3


R.T.: 6.471 min
Delta R.T.: -0.009 min
Response: 36983778
Conc: 437.64 ng/ml



#34 AR-1260-4


R.T.: 7.946 min
Delta R.T.: -0.007 min
Response: 54423957
Conc: 506.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.944 min
Delta R.T.: -0.010 min
Response: 31410435
Conc: 490.78 ng/ml



#35 AR-1260-5

R.T.: 8.262 min
Delta R.T.: -0.007 min
Response: 106785518
Conc: 517.25 ng/ml



#35 AR-1260-5

R.T.: 7.188 min
Delta R.T.: -0.009 min
Response: 66790401
Conc: 493.43 ng/ml