

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022423\
 Data File : PP055969.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Feb 2023 02:11
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 25 07:55:21 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 20 05:24:03 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.321	3.565	86226594	67167302	53.516	45.999
2)	SA Decachlor...	10.061	8.564	95439575	54607739	46.785	48.134
Target Compounds							
16)	L4 AR-1242-1	5.490	4.646	26136453	16978617	541.281	461.463
17)	L4 AR-1242-2	5.513	4.664	37467158	23960043	527.269	454.893
18)	L4 AR-1242-3	5.575	4.840	23888303	12899290	535.441	455.636
19)	L4 AR-1242-4	5.674	4.924	19280915	13655057	533.647	442.586
20)	L4 AR-1242-5	6.413	5.447	21779993	15324355	519.943	460.062



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022423\
Data File : PP055969.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Feb 2023 02:11
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 25 07:55:21 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 20 05:24:03 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.321 min
Delta R.T.: -0.004 min
Response: 86226594
Conc: 53.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.565 min
Delta R.T.: -0.004 min
Response: 67167302
Conc: 46.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.061 min
Delta R.T.: -0.015 min
Response: 95439575
Conc: 46.79 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.564 min
Delta R.T.: -0.013 min
Response: 54607739
Conc: 48.13 ng/ml



#16 AR-1242-1


R.T.: 5.490 min
Delta R.T.: -0.007 min
Response: 26136453
Conc: 541.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.646 min
Delta R.T.: -0.006 min
Response: 16978617
Conc: 461.46 ng/ml




#17 AR-1242-2

R.T.: 5.513 min
Delta R.T.: -0.007 min
Response: 37467158
Conc: 527.27 ng/ml



#17 AR-1242-2


R.T.: 4.664 min
Delta R.T.: -0.006 min
Response: 23960043
Conc: 454.89 ng/ml



#18 AR-1242-3


R.T.: 5.575 min
Delta R.T.: -0.006 min
Response: 23888303
Conc: 535.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.840 min
Delta R.T.: -0.007 min
Response: 12899290
Conc: 455.64 ng/ml




#19 AR-1242-4

R.T.: 5.674 min
Delta R.T.: -0.007 min
Response: 19280915
Conc: 533.65 ng/ml



#19 AR-1242-4


R.T.: 4.924 min
Delta R.T.: -0.006 min
Response: 13655057
Conc: 442.59 ng/ml



#20 AR-1242-5

R.T.: 6.413 min
Delta R.T.: -0.008 min
Response: 21779993
Conc: 519.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.447 min
Delta R.T.: -0.007 min
Response: 15324355
Conc: 460.06 ng/ml