

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022425\
 Data File : PP070027.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Feb 2025 23:22
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP022425AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 25 04:03:03 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 25 04:00:45 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.526	3.830	77660625	50969132	49.934	51.404
2) SA Decachlor...	10.253	8.887	58837660	53434810	48.156	46.645
Target Compounds						
26) L6 AR-1254-1	6.529	5.726	31650110	27652735	522.088	495.991
27) L6 AR-1254-2	6.745	5.874	43660673	24645815	501.928	497.108
28) L6 AR-1254-3	7.107	6.279	44651363	38791765	509.477	491.774
29) L6 AR-1254-4	7.389	6.507	36312560	25589145	494.362	471.693
30) L6 AR-1254-5	7.806	6.926	35741699	34549263	513.376	475.252



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022425\
Data File : PP070027.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Feb 2025 23:22
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP022425AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 25 04:03:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 25 04:00:45 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.526 min
Delta R.T.: -0.001 min
Response: 77660625
Conc: 49.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP022425AR1254




#1 Tetrachloro-m-xylene

R.T.: 3.830 min
Delta R.T.: 0.000 min
Response: 50969132
Conc: 51.40 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.253 min
Delta R.T.: -0.004 min
Response: 58837660
Conc: 48.16 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.887 min
Delta R.T.: -0.001 min
Response: 53434810
Conc: 46.65 ng/ml



#26 AR-1254-1


R.T.: 6.529 min
Delta R.T.: -0.002 min
Response: 31650110
Conc: 522.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP022425AR1254




#26 AR-1254-1

R.T.: 5.726 min
Delta R.T.: -0.002 min
Response: 27652735
Conc: 495.99 ng/ml




#27 AR-1254-2

R.T.: 6.745 min
Delta R.T.: -0.002 min
Response: 43660673
Conc: 501.93 ng/ml



#27 AR-1254-2


R.T.: 5.874 min
Delta R.T.: -0.001 min
Response: 24645815
Conc: 497.11 ng/ml



#28 AR-1254-3


R.T.: 7.107 min
Delta R.T.: -0.002 min
Response: 44651363
Conc: 509.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP022425AR1254




#28 AR-1254-3

R.T.: 6.279 min
Delta R.T.: -0.001 min
Response: 38791765
Conc: 491.77 ng/ml




#29 AR-1254-4

R.T.: 7.389 min
Delta R.T.: -0.003 min
Response: 36312560
Conc: 494.36 ng/ml



#29 AR-1254-4


R.T.: 6.507 min
Delta R.T.: -0.002 min
Response: 25589145
Conc: 471.69 ng/ml



#30 AR-1254-5

R.T.: 7.806 min
Delta R.T.: -0.003 min
Response: 35741699
Conc: 513.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP022425AR1254



#30 AR-1254-5

R.T.: 6.926 min
Delta R.T.: -0.002 min
Response: 34549263
Conc: 475.25 ng/ml