

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022521\
 Data File : PP033573.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 Feb 2021 16:05
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 26 00:20:22 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 22 12:13:48 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.715	3.992	3112536	2350028	42.289	53.363 #
2)	SA Decachlor...	10.528	9.244	2234769	1863201	48.049	58.967
Target Compounds							
3)	L1 AR-1016-1	6.013	5.232	881394	663119	460.504	564.549
4)	L1 AR-1016-2	6.037	5.252	1318163	997739	449.132	559.095
5)	L1 AR-1016-3	6.103	5.442	779992	544370	445.643	565.263 #
6)	L1 AR-1016-4	6.207	5.492	679578	415299	456.082	555.307
7)	L1 AR-1016-5	6.520	5.720	610003	525168	437.441	544.788
31)	L7 AR-1260-1	7.689	6.808	997664	940294	435.473	544.626 #
32)	L7 AR-1260-2	7.953	7.006	1493076	1112427	468.995	565.407
33)	L7 AR-1260-3	8.318	7.159	1300008	1065366	447.997	528.736
34)	L7 AR-1260-4	8.547	7.639	1163242	894843	445.746	552.375
35)	L7 AR-1260-5	8.859	7.888	2817093	2125107	476.964	590.134



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022521\
Data File : PP033573.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 Feb 2021 16:05
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 26 00:20:22 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022121.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 22 12:13:48 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.715 min
Delta R.T.: -0.005 min
Response: 3112536
Conc: 42.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.992 min
Delta R.T.: 0.001 min
Response: 2350028
Conc: 53.36 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.528 min
Delta R.T.: -0.008 min
Response: 2234769
Conc: 48.05 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.244 min
Delta R.T.: -0.005 min
Response: 1863201
Conc: 58.97 ng/ml



#3 AR-1016-1


R.T.: 6.013 min
Delta R.T.: -0.006 min
Response: 881394
Conc: 460.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.232 min
Delta R.T.: -0.003 min
Response: 663119
Conc: 564.55 ng/ml




#4 AR-1016-2

R.T.: 6.037 min
Delta R.T.: -0.006 min
Response: 1318163
Conc: 449.13 ng/ml



#4 AR-1016-2


R.T.: 5.252 min
Delta R.T.: -0.003 min
Response: 997739
Conc: 559.09 ng/ml



#5 AR-1016-3


R.T.: 6.103 min
Delta R.T.: -0.005 min
Response: 779992
Conc: 445.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.442 min
Delta R.T.: -0.003 min
Response: 544370
Conc: 565.26 ng/ml




#6 AR-1016-4

R.T.: 6.207 min
Delta R.T.: -0.007 min
Response: 679578
Conc: 456.08 ng/ml



#6 AR-1016-4


R.T.: 5.492 min
Delta R.T.: -0.003 min
Response: 415299
Conc: 555.31 ng/ml



#7 AR-1016-5


R.T.: 6.520 min
Delta R.T.: -0.005 min
Response: 610003
Conc: 437.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.720 min
Delta R.T.: -0.003 min
Response: 525168
Conc: 544.79 ng/ml




#31 AR-1260-1

R.T.: 7.689 min
Delta R.T.: -0.006 min
Response: 997664
Conc: 435.47 ng/ml



#31 AR-1260-1


R.T.: 6.808 min
Delta R.T.: -0.004 min
Response: 940294
Conc: 544.63 ng/ml



#32 AR-1260-2


R.T.: 7.953 min
Delta R.T.: -0.004 min
Response: 1493076
Conc: 468.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 7.006 min
Delta R.T.: -0.004 min
Response: 1112427
Conc: 565.41 ng/ml




#33 AR-1260-3

R.T.: 8.318 min
Delta R.T.: -0.006 min
Response: 1300008
Conc: 448.00 ng/ml



#33 AR-1260-3


R.T.: 7.159 min
Delta R.T.: -0.004 min
Response: 1065366
Conc: 528.74 ng/ml



#34 AR-1260-4


R.T.: 8.547 min
Delta R.T.: -0.003 min
Response: 1163242
Conc: 445.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.639 min
Delta R.T.: -0.004 min
Response: 894843
Conc: 552.38 ng/ml



#35 AR-1260-5

R.T.: 8.859 min
Delta R.T.: -0.005 min
Response: 2817093
Conc: 476.96 ng/ml



#35 AR-1260-5

R.T.: 7.888 min
Delta R.T.: -0.005 min
Response: 2125107
Conc: 590.13 ng/ml