

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022521\
 Data File : PP033589.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Feb 2021 00:32
 Operator : DD\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 26 01:31:13 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 22 12:13:48 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.720	3.989	2933304	2215193	39.854	50.301 #
2)	SA Decachlor...	10.530	9.244	2204590	1846128	47.400	58.426
Target Compounds							
26)	L6 AR-1254-1	6.919	6.096	822352	995188	421.852	546.587 #
27)	L6 AR-1254-2	7.146	6.255	1350168	855726	436.295	549.875 #
28)	L6 AR-1254-3	7.524	6.674	1409960	1401828	462.054	566.630
29)	L6 AR-1254-4	7.819	6.914	1164478	806663	519.985	635.500
30)	L6 AR-1254-5	8.245	7.339	1178453	1233329	502.538	592.138



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022521\
Data File : PP033589.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Feb 2021 00:32
Operator : DD\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 26 01:31:13 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022121.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 22 12:13:48 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.720 min
Delta R.T.: 0.000 min
Response: 2933304
Conc: 39.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.989 min
Delta R.T.: -0.002 min
Response: 2215193
Conc: 50.30 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.530 min
Delta R.T.: -0.006 min
Response: 2204590
Conc: 47.40 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.244 min
Delta R.T.: -0.006 min
Response: 1846128
Conc: 58.43 ng/ml



#26 AR-1254-1


R.T.: 6.919 min
Delta R.T.: -0.002 min
Response: 822352
Conc: 421.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.096 min
Delta R.T.: -0.004 min
Response: 995188
Conc: 546.59 ng/ml




#27 AR-1254-2

R.T.: 7.146 min
Delta R.T.: -0.002 min
Response: 1350168
Conc: 436.29 ng/ml



#27 AR-1254-2


R.T.: 6.255 min
Delta R.T.: -0.005 min
Response: 855726
Conc: 549.88 ng/ml



#28 AR-1254-3


R.T.: 7.524 min
Delta R.T.: -0.003 min
Response: 1409960
Conc: 462.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.674 min
Delta R.T.: -0.005 min
Response: 1401828
Conc: 566.63 ng/ml




#29 AR-1254-4

R.T.: 7.819 min
Delta R.T.: -0.003 min
Response: 1164478
Conc: 519.98 ng/ml



#29 AR-1254-4


R.T.: 6.914 min
Delta R.T.: -0.005 min
Response: 806663
Conc: 635.50 ng/ml



#30 AR-1254-5

R.T.: 8.245 min
Delta R.T.: -0.004 min
Response: 1178453
Conc: 502.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.339 min
Delta R.T.: -0.005 min
Response: 1233329
Conc: 592.14 ng/ml