

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022522\
 Data File : PP044013.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Feb 2022 15:15
 Operator : YP\AJ
 Sample : N1669-01
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 25 23:43:25 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 11 04:48:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.914	3.170	43080538	39231864	19.484	20.887
2)	SA Decachlor...	10.055	8.541	25174338	32227962	21.193	19.865
Target Compounds							
3)	L1 AR-1016-1	5.175	4.313	82527	39867	1.137	0.651 #
4)	L1 AR-1016-2	5.200	4.336	67386	381491	0.639	4.515 #
5)	L1 AR-1016-3	5.259	4.518	232093	84537	3.525	1.810 #
6)	L1 AR-1016-4	5.368	4.572	92422	31236	1.704	0.836 #
7)	L1 AR-1016-5	5.691	4.800	329397	5436	6.271	0.119 #
8)	L2 AR-1221-1	4.133	3.407	3074889	9971	120.777	0.418 #
9)	L2 AR-1221-2	4.236	3.521f	688401	182932	35.857	10.129 #
10)	L2 AR-1221-3	4.318	3.566	81504	203332	1.419	3.941 #
11)	L3 AR-1232-1	4.318	3.566	81504	203332	1.679	4.726 #
12)	L3 AR-1232-2	0.000	4.336	0	381491	N.D.	10.687 #
13)	L3 AR-1232-3	5.200	4.518	67386	84537	1.565	4.323 #
14)	L3 AR-1232-4	5.368	4.621	92422	23156	4.234	1.339 #
15)	L3 AR-1232-5	5.459	4.800	213641	5436	13.246	0.289 #
16)	L4 AR-1242-1	5.175	4.313	82527	39867	1.469	0.818 #
17)	L4 AR-1242-2	5.200	4.336	67386	381491	0.814	5.666 #
18)	L4 AR-1242-3	5.259	4.518	232093	84537	4.484	2.252 #
19)	L4 AR-1242-4	5.368	4.621	92422	23156	2.164	0.641 #
20)	L4 AR-1242-5	6.174	5.183	386159	40419	8.943	0.939 #
21)	L5 AR-1248-1	5.175	4.313	82527	39867	2.013	1.070 #
22)	L5 AR-1248-2	5.459	4.572	213641	31236	3.760	0.639 #
23)	L5 AR-1248-3	5.691	4.621	329397	23156	4.937	0.451 #
24)	L5 AR-1248-4	6.126	4.800	290788	5436	4.162	0.093 #
25)	L5 AR-1248-5	6.174	5.229	386159	94603	5.586	1.651 #
26)	L6 AR-1254-1	6.094	5.183	291014	40419	3.852	0.463 #
27)	L6 AR-1254-2	6.333	5.339	245221	167708	2.186	2.199
28)	L6 AR-1254-3	6.739	5.771	108444	102074	0.937	0.833
29)	L6 AR-1254-4	7.065	6.016	1872684	40152	23.457	0.509 #
30)	L6 AR-1254-5	7.519	6.472	123568	19072	1.441	0.165 #
31)	L7 AR-1260-1	6.931	5.895f	29190	591414	0.349	7.395 #
32)	L7 AR-1260-2	7.211	6.111	107044	512774	1.183	5.251 #
33)	L7 AR-1260-3	7.600	6.285	169437	33857	2.337	0.371 #
34)	L7 AR-1260-4	7.835	6.802	1150176	73219	13.397	0.905 #
35)	L7 AR-1260-5	8.190	7.069	1216113	45706	7.846	0.247 #
36)	L8 AR-1262-1	7.600	6.521	169437	286309	1.716	2.604 #
37)	L8 AR-1262-2	8.190	6.802	1216113	73219	7.533	0.718 #
38)	L8 AR-1262-3	8.518	7.370	84685	448614	0.774	5.264 #
39)	L8 AR-1262-4	8.642f	7.423f	900995	343553	18.788	2.212 #
40)	L8 AR-1262-5	9.286f	7.980	10251	2654645	0.184	34.124 #
41)	L9 AR-1268-1	8.518	7.370	84685	448614	0.434	1.771 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022522\
 Data File : PP044013.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Feb 2022 15:15
 Operator : YP\AJ
 Sample : N1669-01
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 25 23:43:25 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 11 04:48:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.642f	7.423f	900995	343553	5.147	1.520 #
43) L9	AR-1268-3	8.871	7.661	67249	126413	0.436	0.649 #
44) L9	AR-1268-4	9.286f	7.980	10251	2654645	0.164	29.520 #
45) L9	AR-1268-5	9.720	8.292	84440	152729	0.186	0.254 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022522\
Data File : PP044013.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Feb 2022 15:15
Operator : YP\AJ
Sample : N1669-01
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 25 23:43:25 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 11 04:48:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.914 min
Delta R.T.: -0.006 min
Response: 43080538
Conc: 19.48 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.170 min
Delta R.T.: -0.004 min
Response: 39231864
Conc: 20.89 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.055 min
Delta R.T.: -0.005 min
Response: 25174338
Conc: 21.19 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.541 min
Delta R.T.: -0.009 min
Response: 32227962
Conc: 19.86 ng/ml



#3 AR-1016-1


R.T.: 5.175 min
Delta R.T.: 0.000 min
Response: 82527
Conc: 1.14 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.313 min
Delta R.T.: -0.010 min
Response: 39867
Conc: 0.65 ng/ml




#4 AR-1016-2

R.T.: 5.200 min
Delta R.T.: 0.002 min
Response: 67386
Conc: 0.64 ng/ml



#4 AR-1016-2


R.T.: 4.336 min
Delta R.T.: -0.005 min
Response: 381491
Conc: 4.52 ng/ml



#5 AR-1016-3


R.T.: 5.259 min
Delta R.T.: -0.005 min
Response: 232093
Conc: 3.52 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.518 min
Delta R.T.: -0.010 min
Response: 84537
Conc: 1.81 ng/ml




#6 AR-1016-4

R.T.: 5.368 min
Delta R.T.: -0.002 min
Response: 92422
Conc: 1.70 ng/ml



#6 AR-1016-4


R.T.: 4.572 min
Delta R.T.: -0.004 min
Response: 31236
Conc: 0.84 ng/ml



#7 AR-1016-5


R.T.: 5.691 min
Delta R.T.: 0.000 min
Response: 329397
Conc: 6.27 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 4.800 min
Delta R.T.: -0.002 min
Response: 5436
Conc: 0.12 ng/ml




#8 AR-1221-1

R.T.: 4.133 min
Delta R.T.: -0.014 min
Response: 3074889
Conc: 120.78 ng/ml



#8 AR-1221-1


R.T.: 3.407 min
Delta R.T.: 0.004 min
Response: 9971
Conc: 0.42 ng/ml



#9 AR-1221-2


R.T.: 4.236 min
Delta R.T.: -0.002 min
Response: 688401
Conc: 35.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.521 min
Delta R.T.: 0.029 min
Response: 182932
Conc: 10.13 ng/ml




#10 AR-1221-3

R.T.: 4.318 min
Delta R.T.: 0.000 min
Response: 81504
Conc: 1.42 ng/ml



#10 AR-1221-3


R.T.: 3.566 min
Delta R.T.: -0.006 min
Response: 203332
Conc: 3.94 ng/ml



#11 AR-1232-1


R.T.: 4.318 min
Delta R.T.: 0.000 min
Response: 81504
Conc: 1.68 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.566 min
Delta R.T.: -0.006 min
Response: 203332
Conc: 4.73 ng/ml




#12 AR-1232-2

R.T.: 0.000 min
Exp R.T. : 4.884 min
Response: 0
Conc: N.D.



#12 AR-1232-2


R.T.: 4.336 min
Delta R.T.: -0.005 min
Response: 381491
Conc: 10.69 ng/ml



#13 AR-1232-3


R.T.: 5.200 min
Delta R.T.: 0.003 min
Response: 67386
Conc: 1.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.518 min
Delta R.T.: -0.009 min
Response: 84537
Conc: 4.32 ng/ml




#14 AR-1232-4

R.T.: 5.368 min
Delta R.T.: -0.001 min
Response: 92422
Conc: 4.23 ng/ml



#14 AR-1232-4


R.T.: 4.621 min
Delta R.T.: 0.002 min
Response: 23156
Conc: 1.34 ng/ml



#15 AR-1232-5


R.T.: 5.459 min
Delta R.T.: -0.013 min
Response: 213641
Conc: 13.25 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 4.800 min
Delta R.T.: -0.001 min
Response: 5436
Conc: 0.29 ng/ml




#16 AR-1242-1

R.T.: 5.175 min
Delta R.T.: 0.000 min
Response: 82527
Conc: 1.47 ng/ml



#16 AR-1242-1


R.T.: 4.313 min
Delta R.T.: -0.010 min
Response: 39867
Conc: 0.82 ng/ml



#17 AR-1242-2


R.T.: 5.200 min
Delta R.T.: 0.003 min
Response: 67386
Conc: 0.81 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.336 min
Delta R.T.: -0.006 min
Response: 381491
Conc: 5.67 ng/ml




#18 AR-1242-3

R.T.: 5.259 min
Delta R.T.: -0.005 min
Response: 232093
Conc: 4.48 ng/ml



#18 AR-1242-3


R.T.: 4.518 min
Delta R.T.: -0.009 min
Response: 84537
Conc: 2.25 ng/ml



#19 AR-1242-4


R.T.: 5.368 min
Delta R.T.: -0.002 min
Response: 92422
Conc: 2.16 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.621 min
Delta R.T.: 0.002 min
Response: 23156
Conc: 0.64 ng/ml




#20 AR-1242-5

R.T.: 6.174 min
Delta R.T.: 0.003 min
Response: 386159
Conc: 8.94 ng/ml



#20 AR-1242-5


R.T.: 5.183 min
Delta R.T.: 0.004 min
Response: 40419
Conc: 0.94 ng/ml



#21 AR-1248-1


R.T.: 5.175 min
Delta R.T.: 0.000 min
Response: 82527
Conc: 2.01 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.313 min
Delta R.T.: -0.010 min
Response: 39867
Conc: 1.07 ng/ml




#22 AR-1248-2

R.T.: 5.459 min
Delta R.T.: -0.013 min
Response: 213641
Conc: 3.76 ng/ml



#22 AR-1248-2


R.T.: 4.572 min
Delta R.T.: -0.004 min
Response: 31236
Conc: 0.64 ng/ml



#23 AR-1248-3


R.T.: 5.691 min
Delta R.T.: 0.001 min
Response: 329397
Conc: 4.94 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.621 min
Delta R.T.: 0.002 min
Response: 23156
Conc: 0.45 ng/ml




#24 AR-1248-4

R.T.: 6.126 min
Delta R.T.: -0.004 min
Response: 290788
Conc: 4.16 ng/ml



#24 AR-1248-4


R.T.: 4.800 min
Delta R.T.: -0.002 min
Response: 5436
Conc: 0.09 ng/ml



#25 AR-1248-5


R.T.: 6.174 min
Delta R.T.: 0.003 min
Response: 386159
Conc: 5.59 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.229 min
Delta R.T.: 0.006 min
Response: 94603
Conc: 1.65 ng/ml




#26 AR-1254-1

R.T.: 6.094 min
Delta R.T.: -0.008 min
Response: 291014
Conc: 3.85 ng/ml



#26 AR-1254-1


R.T.: 5.183 min
Delta R.T.: 0.004 min
Response: 40419
Conc: 0.46 ng/ml



#27 AR-1254-2


R.T.: 6.333 min
Delta R.T.: -0.008 min
Response: 245221
Conc: 2.19 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.339 min
Delta R.T.: 0.000 min
Response: 167708
Conc: 2.20 ng/ml




#28 AR-1254-3

R.T.: 6.739 min
Delta R.T.: -0.003 min
Response: 108444
Conc: 0.94 ng/ml



#28 AR-1254-3


R.T.: 5.771 min
Delta R.T.: 0.000 min
Response: 102074
Conc: 0.83 ng/ml



#29 AR-1254-4


R.T.: 7.065 min
Delta R.T.: 0.009 min
Response: 1872684
Conc: 23.46 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.016 min
Delta R.T.: -0.006 min
Response: 40152
Conc: 0.51 ng/ml




#30 AR-1254-5

R.T.: 7.519 min
Delta R.T.: 0.002 min
Response: 123568
Conc: 1.44 ng/ml



#30 AR-1254-5


R.T.: 6.472 min
Delta R.T.: -0.002 min
Response: 19072
Conc: 0.16 ng/ml



#31 AR-1260-1


R.T.: 6.931 min
Delta R.T.: 0.009 min
Response: 29190
Conc: 0.35 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 5.895 min
Delta R.T.: -0.018 min
Response: 591414
Conc: 7.40 ng/ml




#32 AR-1260-2

R.T.: 7.211 min
Delta R.T.: 0.005 min
Response: 107044
Conc: 1.18 ng/ml



#32 AR-1260-2


R.T.: 6.111 min
Delta R.T.: -0.009 min
Response: 512774
Conc: 5.25 ng/ml



#33 AR-1260-3


R.T.: 7.600 min
Delta R.T.: 0.000 min
Response: 169437
Conc: 2.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.285 min
Delta R.T.: 0.002 min
Response: 33857
Conc: 0.37 ng/ml




#34 AR-1260-4

R.T.: 7.835 min
Delta R.T.: -0.014 min
Response: 1150176
Conc: 13.40 ng/ml



#34 AR-1260-4


R.T.: 6.802 min
Delta R.T.: 0.006 min
Response: 73219
Conc: 0.90 ng/ml



#35 AR-1260-5


R.T.: 8.190 min
Delta R.T.: -0.002 min
Response: 1216113
Conc: 7.85 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.069 min
Delta R.T.: 0.005 min
Response: 45706
Conc: 0.25 ng/ml




#36 AR-1262-1

R.T.: 7.600 min
Delta R.T.: -0.002 min
Response: 169437
Conc: 1.72 ng/ml



#36 AR-1262-1


R.T.: 6.521 min
Delta R.T.: 0.004 min
Response: 286309
Conc: 2.60 ng/ml



#37 AR-1262-2


R.T.: 8.190 min
Delta R.T.: -0.003 min
Response: 1216113
Conc: 7.53 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.802 min
Delta R.T.: 0.005 min
Response: 73219
Conc: 0.72 ng/ml




#38 AR-1262-3

R.T.: 8.518 min
Delta R.T.: -0.008 min
Response: 84685
Conc: 0.77 ng/ml



#38 AR-1262-3


R.T.: 7.370 min
Delta R.T.: -0.002 min
Response: 448614
Conc: 5.26 ng/ml



#39 AR-1262-4


R.T.: 8.642 min
Delta R.T.: 0.023 min
Response: 900995
Conc: 18.79 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.423 min
Delta R.T.: -0.018 min
Response: 343553
Conc: 2.21 ng/ml




#40 AR-1262-5

R.T.: 9.286 min
Delta R.T.: -0.017 min
Response: 10251
Conc: 0.18 ng/ml



#40 AR-1262-5


R.T.: 7.980 min
Delta R.T.: -0.004 min
Response: 2654645
Conc: 34.12 ng/ml



#41 AR-1268-1


R.T.: 8.518 min
Delta R.T.: -0.004 min
Response: 84685
Conc: 0.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.370 min
Delta R.T.: -0.001 min
Response: 448614
Conc: 1.77 ng/ml




#42 AR-1268-2

R.T.: 8.642 min
Delta R.T.: 0.021 min
Response: 900995
Conc: 5.15 ng/ml



#42 AR-1268-2


R.T.: 7.423 min
Delta R.T.: -0.019 min
Response: 343553
Conc: 1.52 ng/ml



#43 AR-1268-3


R.T.: 8.871 min
Delta R.T.: 0.012 min
Response: 67249
Conc: 0.44 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.661 min
Delta R.T.: -0.004 min
Response: 126413
Conc: 0.65 ng/ml




#44 AR-1268-4

R.T.: 9.286 min
Delta R.T.: -0.016 min
Response: 10251
Conc: 0.16 ng/ml



#44 AR-1268-4


R.T.: 7.980 min
Delta R.T.: -0.004 min
Response: 2654645
Conc: 29.52 ng/ml



#45 AR-1268-5

R.T.: 9.720 min
Delta R.T.: -0.008 min
Response: 84440
Conc: 0.19 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.292 min
Delta R.T.: 0.001 min
Response: 152729
Conc: 0.25 ng/ml