

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022522\
 Data File : PP043999.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Feb 2022 11:40
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 25 23:33:24 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 11 04:48:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.914	3.171	111.4E6	95691274	50.404	50.947
2) SA Decachlor...	10.056	8.542	65408584	79651708	55.063	49.095
Target Compounds						
16) L4 AR-1242-1	5.168	4.316	26592597	23527170	473.247	482.629
17) L4 AR-1242-2	5.191	4.335	38894913	33076554	470.120	491.224
18) L4 AR-1242-3	5.257	4.520	24213933	18523188	467.812	493.445
19) L4 AR-1242-4	5.362	4.612	20173597	17827251	472.312	493.700
20) L4 AR-1242-5	6.164	5.171	20061848	20883695	464.589	485.126



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022522\
Data File : PP043999.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Feb 2022 11:40
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 25 23:33:24 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 11 04:48:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.914 min
Delta R.T.: -0.005 min
Response: 111444813
Conc: 50.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.171 min
Delta R.T.: -0.004 min
Response: 95691274
Conc: 50.95 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.056 min
Delta R.T.: -0.004 min
Response: 65408584
Conc: 55.06 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.542 min
Delta R.T.: -0.008 min
Response: 79651708
Conc: 49.10 ng/ml



#16 AR-1242-1


R.T.: 5.168 min
Delta R.T.: -0.007 min
Response: 26592597
Conc: 473.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.316 min
Delta R.T.: -0.007 min
Response: 23527170
Conc: 482.63 ng/ml




#17 AR-1242-2

R.T.: 5.191 min
Delta R.T.: -0.006 min
Response: 38894913
Conc: 470.12 ng/ml



#17 AR-1242-2


R.T.: 4.335 min
Delta R.T.: -0.007 min
Response: 33076554
Conc: 491.22 ng/ml



#18 AR-1242-3


R.T.: 5.257 min
Delta R.T.: -0.007 min
Response: 24213933
Conc: 467.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.520 min
Delta R.T.: -0.007 min
Response: 18523188
Conc: 493.45 ng/ml




#19 AR-1242-4

R.T.: 5.362 min
Delta R.T.: -0.008 min
Response: 20173597
Conc: 472.31 ng/ml



#19 AR-1242-4


R.T.: 4.612 min
Delta R.T.: -0.007 min
Response: 17827251
Conc: 493.70 ng/ml



#20 AR-1242-5

R.T.: 6.164 min
Delta R.T.: -0.007 min
Response: 20061848
Conc: 464.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.171 min
Delta R.T.: -0.008 min
Response: 20883695
Conc: 485.13 ng/ml