

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022621\
 Data File : PP033594.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Feb 2021 11:05
 Operator : DD\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 26 14:00:04 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 22 12:13:48 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.727	3.990	3123136	2000598	42.433	45.428
2) SA Decachlor...	10.535	9.245	2313341	1707149	49.738	54.028
Target Compounds						
26) L6 AR-1254-1	6.925	6.100	793668	889709	407.138	488.655
27) L6 AR-1254-2	7.152	6.258	1426067	766592	460.821	492.600
28) L6 AR-1254-3	7.530	6.676	1529018	1291536	501.070	522.050
29) L6 AR-1254-4	7.824	6.916	1315483	748098	587.414	589.361
30) L6 AR-1254-5	8.250	7.341	1284169	1132475	547.619	543.717



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022621\
Data File : PP033594.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Feb 2021 11:05
Operator : DD\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 26 14:00:04 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022121.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 22 12:13:48 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.727 min
Delta R.T.: 0.007 min
Response: 3123136
Conc: 42.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.990 min
Delta R.T.: 0.000 min
Response: 2000598
Conc: 45.43 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.535 min
Delta R.T.: 0.000 min
Response: 2313341
Conc: 49.74 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.245 min
Delta R.T.: -0.005 min
Response: 1707149
Conc: 54.03 ng/ml



#26 AR-1254-1


R.T.: 6.925 min
Delta R.T.: 0.004 min
Response: 793668
Conc: 407.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.100 min
Delta R.T.: 0.000 min
Response: 889709
Conc: 488.65 ng/ml




#27 AR-1254-2

R.T.: 7.152 min
Delta R.T.: 0.004 min
Response: 1426067
Conc: 460.82 ng/ml



#27 AR-1254-2


R.T.: 6.258 min
Delta R.T.: -0.002 min
Response: 766592
Conc: 492.60 ng/ml



#28 AR-1254-3


R.T.: 7.530 min
Delta R.T.: 0.003 min
Response: 1529018
Conc: 501.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.676 min
Delta R.T.: -0.003 min
Response: 1291536
Conc: 522.05 ng/ml




#29 AR-1254-4

R.T.: 7.824 min
Delta R.T.: 0.002 min
Response: 1315483
Conc: 587.41 ng/ml



#29 AR-1254-4


R.T.: 6.916 min
Delta R.T.: -0.002 min
Response: 748098
Conc: 589.36 ng/ml



#30 AR-1254-5

R.T.: 8.250 min
Delta R.T.: 0.000 min
Response: 1284169
Conc: 547.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.341 min
Delta R.T.: -0.002 min
Response: 1132475
Conc: 543.72 ng/ml