

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022624\
 Data File : PP063437.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Feb 2024 20:43
 Operator : YP\AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 27 00:52:24 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 27 00:49:16 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.489	3.640	57383791	58788072	25.144	24.876
2) SA Decachlor...	10.361	8.684	90629689	90309981	26.001	25.759
Target Compounds						
41) L9 AR-1268-1	8.793	7.572	64556134	78034959	256.534	253.445
42) L9 AR-1268-2	8.890	7.637	56532428	72101337	255.308	253.834
43) L9 AR-1268-3	9.128	7.846	54940356	67693544	256.671	254.646
44) L9 AR-1268-4	9.570	8.133	18157805	20045667	248.146	260.217
45) L9 AR-1268-5	10.005	8.426	160.0E6	148.8E6	252.192	246.288



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022624\
Data File : PP063437.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Feb 2024 20:43
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 27 00:52:24 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022624.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 27 00:49:16 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.489 min
Delta R.T.: 0.000 min
Response: 57383791
Conc: 25.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.640 min
Delta R.T.: 0.000 min
Response: 58788072
Conc: 24.88 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.361 min
Delta R.T.: 0.000 min
Response: 90629689
Conc: 26.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.684 min
Delta R.T.: 0.000 min
Response: 90309981
Conc: 25.76 ng/ml



#41 AR-1268-1


R.T.: 8.793 min
Delta R.T.: 0.000 min
Response: 64556134
Conc: 256.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#41 AR-1268-1

R.T.: 7.572 min
Delta R.T.: 0.000 min
Response: 78034959
Conc: 253.45 ng/ml




#42 AR-1268-2

R.T.: 8.890 min
Delta R.T.: -0.001 min
Response: 56532428
Conc: 255.31 ng/ml



#42 AR-1268-2


R.T.: 7.637 min
Delta R.T.: 0.000 min
Response: 72101337
Conc: 253.83 ng/ml



#43 AR-1268-3


R.T.: 9.128 min
Delta R.T.: 0.000 min
Response: 54940356
Conc: 256.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#43 AR-1268-3

R.T.: 7.846 min
Delta R.T.: 0.000 min
Response: 67693544
Conc: 254.65 ng/ml




#44 AR-1268-4

R.T.: 9.570 min
Delta R.T.: 0.000 min
Response: 18157805
Conc: 248.15 ng/ml



#44 AR-1268-4


R.T.: 8.133 min
Delta R.T.: 0.000 min
Response: 20045667
Conc: 260.22 ng/ml



#45 AR-1268-5

R.T.: 10.005 min
Delta R.T.: 0.000 min
Response: 159996777
Conc: 252.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250



#45 AR-1268-5

R.T.: 8.426 min
Delta R.T.: 0.000 min
Response: 148826385
Conc: 246.29 ng/ml