

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022719\
 Data File : PP029182.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Feb 2019 19:58
 Operator : EI\MA
 Sample : K1348-08MSD
 Misc : hexane
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 OK-02-022519MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 28 03:37:19 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022119.M
 Quant Title : 8080.M
 QLast Update : Fri Feb 22 02:24:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.937	12.126	17875673	20307946	594.738	535.088
Target Compounds							
1) T	Dalapon	3.245	2.957	33431555	28683635	570.713	554.292
2) T	3,5-DICHL...	10.346	10.213	36228020	29566602	587.309	583.364
5) T	DICAMBA	12.259	12.439	79517036	77616486	598.013	606.378
8) T	DICHLORPROP	13.448	13.550	20577566	23238459	666.899	687.350
9) T	2,4-D	13.797	14.036	25434360	30664687	648.716	676.274
10) T	Pentachlo...	14.214	14.643	357.8E6	289.6E6	549.127	562.687
11) T	2,4,5-TP ...	15.072	15.214	141.9E6	124.4E6	598.273	611.415
12) T	2,4,5-T	15.460	15.862	136.1E6	99578220	644.500	601.417
13) T	2,4-DB	16.218	16.457	28226493	18947088	1477.647	841.099 #
14) T	DINOSEB	17.682	16.848	76426453	74084458	524.797	565.219
15) T	Picloram	17.472	18.243	179.7E6	148.6E6	646.427	660.288
16) T	DCPA	18.066	18.117	116.1E6	106.5E6	410.968	458.202



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022719\
Data File : PP029182.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 Feb 2019 19:58
Operator : EI\MA
Sample : K1348-08MSD
Misc : hexane
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
OK-02-022519MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 28 03:37:19 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022119.M
Quant Title : 8080.M
QLast Update : Fri Feb 22 02:24:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.245 min
Delta R.T.: 0.000 min
Response: 33431555
Conc: 570.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
OK-02-022519MSD




#1 Dalapon

R.T.: 2.957 min
Delta R.T.: -0.001 min
Response: 28683635
Conc: 554.29 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 10.346 min
Delta R.T.: -0.002 min
Response: 36228020
Conc: 587.31 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 10.213 min
Delta R.T.: -0.002 min
Response: 29566602
Conc: 583.36 ng/ml



#4 2,4-DCAA


R.T.: 11.937 min
Delta R.T.: -0.004 min
Response: 17875673
Conc: 594.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
OK-02-022519MSD




#4 2,4-DCAA

R.T.: 12.126 min
Delta R.T.: -0.002 min
Response: 20307946
Conc: 535.09 ng/ml




#5 DICAMBA

R.T.: 12.259 min
Delta R.T.: -0.004 min
Response: 79517036
Conc: 598.01 ng/ml



#5 DICAMBA


R.T.: 12.439 min
Delta R.T.: -0.002 min
Response: 77616486
Conc: 606.38 ng/ml



#8 DICHLORPROP


R.T.: 13.448 min
Delta R.T.: -0.004 min
Response: 20577566
Conc: 666.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
OK-02-022519MSD




#8 DICHLORPROP

R.T.: 13.550 min
Delta R.T.: -0.002 min
Response: 23238459
Conc: 687.35 ng/ml




#9 2,4-D

R.T.: 13.797 min
Delta R.T.: -0.003 min
Response: 25434360
Conc: 648.72 ng/ml



#9 2,4-D


R.T.: 14.036 min
Delta R.T.: -0.002 min
Response: 30664687
Conc: 676.27 ng/ml



#10 Pentachlorophenol


R.T.: 14.214 min
Delta R.T.: -0.002 min
Response: 357829263
Conc: 549.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
OK-02-022519MSD




#10 Pentachlorophenol

R.T.: 14.643 min
Delta R.T.: -0.002 min
Response: 289596035
Conc: 562.69 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 15.072 min
Delta R.T.: -0.003 min
Response: 141935494
Conc: 598.27 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 15.214 min
Delta R.T.: -0.001 min
Response: 124365865
Conc: 611.41 ng/ml



#12 2,4,5-T


R.T.: 15.460 min
Delta R.T.: -0.004 min
Response: 136063146
Conc: 644.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
OK-02-022519MSD




#12 2,4,5-T

R.T.: 15.862 min
Delta R.T.: -0.002 min
Response: 99578220
Conc: 601.42 ng/ml




#13 2,4-DB

R.T.: 16.218 min
Delta R.T.: -0.005 min
Response: 28226493
Conc: 1477.65 ng/ml



#13 2,4-DB


R.T.: 16.457 min
Delta R.T.: -0.002 min
Response: 18947088
Conc: 841.10 ng/ml



#14 DINOSEB


R.T.: 17.682 min
Delta R.T.: -0.004 min
Response: 76426453
Conc: 524.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
OK-02-022519MSD




#14 DINOSEB

R.T.: 16.848 min
Delta R.T.: -0.002 min
Response: 74084458
Conc: 565.22 ng/ml




#15 Picloram

R.T.: 17.472 min
Delta R.T.: -0.004 min
Response: 179738976
Conc: 646.43 ng/ml



#15 Picloram


R.T.: 18.243 min
Delta R.T.: -0.002 min
Response: 148579925
Conc: 660.29 ng/ml



#16 DCPA

R.T.: 18.066 min
Delta R.T.: -0.003 min
Response: 116089227
Conc: 410.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
OK-02-022519MSD



#16 DCPA

R.T.: 18.117 min
Delta R.T.: -0.001 min
Response: 106484461
Conc: 458.20 ng/ml