

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022719\
 Data File : PP029185.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Feb 2019 21:15
 Operator : EI\MA
 Sample : HSTDCCC750
 Misc : hexane
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 28 03:37:48 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022119.M
 Quant Title : 8080.M
 QLast Update : Fri Feb 22 02:24:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.937	12.127	20789843	28703451	691.695	756.299
Target Compounds							
1) T	Dalapon	3.245	2.957	40152859	35022742	685.453	676.791
2) T	3,5-DICHL...	10.346	10.214	42006200	35542857	680.981	701.278
3) T	4-Nitroph...	11.500	11.272	47988726	16041182	700.527	669.839
5) T	DICAMBA	12.258	12.438	92937854	91385876	698.945	713.951
8) T	DICHLORPROP	13.449	13.550	22176983	24615459	718.735	728.079
9) T	2,4-D	13.796	14.036	28045431	32636391	715.312	719.758
10) T	Pentachlo...	14.214	14.643	457.9E6	372.8E6	702.651	724.284
11) T	2,4,5-TP ...	15.073	15.214	166.7E6	146.1E6	702.682	718.219
12) T	2,4,5-T	15.461	15.863	149.7E6	117.5E6	708.892	709.524
13) T	2,4-DB	16.218	16.458	13603145	16390555	712.120	727.609
14) T	DINOSEB	17.682	16.847	100.1E6	90421454	687.574	689.860
15) T	Picloram	17.472	18.243	199.2E6	162.1E6	716.476	720.198
16) T	DCPA	18.067	18.117	201.3E6	168.4E6	712.495	724.515

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022719\
Data File : PP029185.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 Feb 2019 21:15
Operator : EI\MA
Sample : HSTDCCC750
Misc : hexane
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 28 03:37:48 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022119.M
Quant Title : 8080.M
QLast Update : Fri Feb 22 02:24:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

