

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022725\
 Data File : PP070081.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Feb 2025 09:44
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 27 14:59:32 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 25 05:10:19 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.526	3.830	77200828	49477820	52.603	51.753
2) SA Decachlor...	10.256	8.887	54320615	50184453	47.690	46.291
Target Compounds						
21) L5 AR-1248-1	5.679	4.918	16780162	10862259	512.995	513.104m
22) L5 AR-1248-2	5.951	5.157	22727916	14803935	539.187	523.988
23) L5 AR-1248-3	6.154	5.199	23022532	15432063	509.661	518.627
24) L5 AR-1248-4	6.552	5.372	28443732	17965189	482.397	511.168
25) L5 AR-1248-5	6.591	5.767	27143381	18277010	474.650	505.336



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022725\
Data File : PP070081.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Feb 2025 09:44
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 27 14:59:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 25 05:10:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.526 min
Delta R.T.: 0.000 min
Response: 77200828
Conc: 52.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.830 min
Delta R.T.: 0.000 min
Response: 49477820
Conc: 51.75 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.256 min
Delta R.T.: 0.000 min
Response: 54320615
Conc: 47.69 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.887 min
Delta R.T.: 0.000 min
Response: 50184453
Conc: 46.29 ng/ml



#21 AR-1248-1


R.T.: 5.679 min
Delta R.T.: -0.002 min
Response: 16780162
Conc: 513.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.918 min
Delta R.T.: -0.021 min
Response: 10862259
Conc: 513.10 ng/ml




#22 AR-1248-2

R.T.: 5.951 min
Delta R.T.: -0.002 min
Response: 22727916
Conc: 539.19 ng/ml



#22 AR-1248-2


R.T.: 5.157 min
Delta R.T.: -0.001 min
Response: 14803935
Conc: 523.99 ng/ml



#23 AR-1248-3


R.T.: 6.154 min
Delta R.T.: -0.002 min
Response: 23022532
Conc: 509.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.199 min
Delta R.T.: -0.001 min
Response: 15432063
Conc: 518.63 ng/ml




#24 AR-1248-4

R.T.: 6.552 min
Delta R.T.: 0.000 min
Response: 28443732
Conc: 482.40 ng/ml



#24 AR-1248-4


R.T.: 5.372 min
Delta R.T.: -0.001 min
Response: 17965189
Conc: 511.17 ng/ml



#25 AR-1248-5

R.T.: 6.591 min
Delta R.T.: -0.003 min
Response: 27143381
Conc: 474.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.767 min
Delta R.T.: 0.000 min
Response: 18277010
Conc: 505.34 ng/ml