

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022725\
 Data File : PP070097.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Feb 2025 17:27
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/28/2025
 Supervised By :Ankita Jodhani 02/28/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 28 00:21:00 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 25 05:10:19 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.530	3.832	75551031	48516211	51.478	50.747
2) SA Decachlor...	10.261	8.891	54407002	48356047	47.766	44.605
Target Compounds						
26) L6 AR-1254-1	6.533	5.728	29186233	26587946	514.745	487.232
27) L6 AR-1254-2	6.749	5.877	42275275	23618076	482.124	484.518
28) L6 AR-1254-3	7.113	6.282	41323806	36447806	463.654	480.737
29) L6 AR-1254-4	7.395	6.510	36531541	24584297	490.283	477.736m
30) L6 AR-1254-5	7.812	6.929	31311222	33455838	466.202	495.134

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022725\
Data File : PP070097.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Feb 2025 17:27
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/28/2025
Supervised By :Ankita Jodhani 02/28/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 28 00:21:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022425.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 25 05:10:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.530 min
Delta R.T.: 0.003 min
Response: 75551031
Conc: 51.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/28/2025
Supervised By :Ankita Jodhani 02/28/2025




#1 Tetrachloro-m-xylene

R.T.: 3.832 min
Delta R.T.: 0.002 min
Response: 48516211
Conc: 50.75 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.261 min
Delta R.T.: 0.005 min
Response: 54407002
Conc: 47.77 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.891 min
Delta R.T.: 0.003 min
Response: 48356047
Conc: 44.60 ng/ml




#26 AR-1254-1

R.T.: 6.533 min
Delta R.T.: 0.003 min
Response: 29186233
Conc: 514.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/28/2025
Supervised By :Ankita Jodhani 02/28/2025




#26 AR-1254-1

R.T.: 5.728 min
Delta R.T.: 0.002 min
Response: 26587946
Conc: 487.23 ng/ml




#27 AR-1254-2

R.T.: 6.749 min
Delta R.T.: 0.002 min
Response: 42275275
Conc: 482.12 ng/ml



#27 AR-1254-2

R.T.: 5.877 min
Delta R.T.: 0.002 min
Response: 23618076
Conc: 484.52 ng/ml




#28 AR-1254-3

R.T.: 7.113 min
Delta R.T.: 0.003 min
Response: 41323806
Conc: 463.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/28/2025
Supervised By :Ankita Jodhani 02/28/2025




#28 AR-1254-3

R.T.: 6.282 min
Delta R.T.: -0.014 min
Response: 36447806
Conc: 480.74 ng/ml




#29 AR-1254-4

R.T.: 7.395 min
Delta R.T.: 0.002 min
Response: 36531541
Conc: 490.28 ng/ml



#29 AR-1254-4

R.T.: 6.510 min
Delta R.T.: -0.029 min
Response: 24584297
Conc: 477.74 ng/ml m




#30 AR-1254-5

R.T.: 7.812 min
Delta R.T.: 0.003 min
Response: 31311222
Conc: 466.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/28/2025
Supervised By :Ankita Jodhani 02/28/2025



#30 AR-1254-5

R.T.: 6.929 min
Delta R.T.: 0.002 min
Response: 33455838
Conc: 495.13 ng/ml