

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022822\
 Data File : PP044075.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Feb 2022 13:58
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 01 00:35:10 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 11 04:48:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.912	3.168	109.2E6	91110371	49.395	48.508
2) SA Decachlor...	10.055	8.540	62941728	73582370	52.986	45.354
Target Compounds						
16) L4 AR-1242-1	5.166	4.314	26376377	22516114	469.399	461.889
17) L4 AR-1242-2	5.189	4.333	38342345	31226439	463.441	463.748
18) L4 AR-1242-3	5.255	4.518	23935672	17613216	462.436	469.204
19) L4 AR-1242-4	5.361	4.610	19860525	16929402	464.982	468.835
20) L4 AR-1242-5	6.162	5.169	19909894	19801855	461.070	459.995



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022822\
Data File : PP044075.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Feb 2022 13:58
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 01 00:35:10 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021022.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 11 04:48:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.912 min
Delta R.T.: -0.007 min
Response: 109214237
Conc: 49.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.168 min
Delta R.T.: -0.006 min
Response: 91110371
Conc: 48.51 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.055 min
Delta R.T.: -0.005 min
Response: 62941728
Conc: 52.99 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.540 min
Delta R.T.: -0.010 min
Response: 73582370
Conc: 45.35 ng/ml



#16 AR-1242-1


R.T.: 5.166 min
Delta R.T.: -0.008 min
Response: 26376377
Conc: 469.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.314 min
Delta R.T.: -0.009 min
Response: 22516114
Conc: 461.89 ng/ml




#17 AR-1242-2

R.T.: 5.189 min
Delta R.T.: -0.008 min
Response: 38342345
Conc: 463.44 ng/ml



#17 AR-1242-2


R.T.: 4.333 min
Delta R.T.: -0.009 min
Response: 31226439
Conc: 463.75 ng/ml



#18 AR-1242-3


R.T.: 5.255 min
Delta R.T.: -0.008 min
Response: 23935672
Conc: 462.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.518 min
Delta R.T.: -0.009 min
Response: 17613216
Conc: 469.20 ng/ml




#19 AR-1242-4

R.T.: 5.361 min
Delta R.T.: -0.008 min
Response: 19860525
Conc: 464.98 ng/ml



#19 AR-1242-4


R.T.: 4.610 min
Delta R.T.: -0.009 min
Response: 16929402
Conc: 468.84 ng/ml



#20 AR-1242-5

R.T.: 6.162 min
Delta R.T.: -0.009 min
Response: 19909894
Conc: 461.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.169 min
Delta R.T.: -0.010 min
Response: 19801855
Conc: 460.00 ng/ml