

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022823\
 Data File : PP056046.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Feb 2023 10:39
 Operator : YP\AJ
 Sample : PB151053BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 28 22:08:10 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 20 05:24:03 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.323	3.566	31464301	23302384	19.528	15.959
2)	SA Decachlor...	10.072	8.570	35102355	22975079	17.207	20.251
Target Compounds							
3)	L1 AR-1016-1	5.493	0.000	8456	0	0.142	N.D. #
4)	L1 AR-1016-2	5.523	0.000	7750	0	0.089	N.D. #
5)	L1 AR-1016-3	5.577	4.848	62410	84536	1.123	2.432 #
6)	L1 AR-1016-4	5.677	0.000	36181	0	0.812	N.D. #
7)	L1 AR-1016-5	5.977	0.000	19166	0	0.397	N.D. #
8)	L2 AR-1221-1	4.537	0.000	46018	0	2.211	N.D. #
9)	L2 AR-1221-2	4.629	3.864	524485	375009	32.973	28.554
10)	L2 AR-1221-3	4.694	3.953	24482	44626	0.505	1.069 #
11)	L3 AR-1232-1	4.694	3.953	24482	44626	0.618	1.296 #
12)	L3 AR-1232-2	5.230	0.000	9509	0	0.458	N.D. #
13)	L3 AR-1232-3	5.523	4.848	7750	84536	0.197	5.563 #
14)	L3 AR-1232-4	5.677	4.947f	36181	230789	1.847	15.224 #
15)	L3 AR-1232-5	5.777	0.000	40997	0	2.563	N.D. #
16)	L4 AR-1242-1	5.493	0.000	8456	0	0.175	N.D. #
17)	L4 AR-1242-2	5.523	0.000	7750	0	0.109	N.D. #
18)	L4 AR-1242-3	5.577	4.848	62410	84536	1.399	2.986 #
19)	L4 AR-1242-4	5.677	4.947f	36181	230789	1.001	7.480 #
20)	L4 AR-1242-5	6.421	5.474f	47736	844868	1.140	25.364 #
21)	L5 AR-1248-1	5.493	0.000	8456	0	0.232	N.D. #
22)	L5 AR-1248-2	5.777	0.000	40997	0	0.713	N.D. #
23)	L5 AR-1248-3	5.977	4.947f	19166	230789	0.295	5.186 #
24)	L5 AR-1248-4	6.389	0.000	219221	0	3.022	N.D. #
25)	L5 AR-1248-5	6.421	5.474f	47736	844868	0.673	19.168 #
26)	L6 AR-1254-1	6.356	5.474f	750054	844868	10.202	11.484
27)	L6 AR-1254-2	6.576	5.605	59565	320437	0.516	5.093 #
28)	L6 AR-1254-3	6.941	6.033f	293103	707267	2.382	7.211 #
29)	L6 AR-1254-4	7.217	0.000	76793	0	0.836	N.D. #
30)	L6 AR-1254-5	7.655	0.000	504101	0	4.796	N.D. #
31)	L7 AR-1260-1	7.105	6.132	121016	159379	1.169	2.186 #
32)	L7 AR-1260-2	7.370	6.307f	95282	265105	0.758	3.249 #
33)	L7 AR-1260-3	7.718	0.000	33395	0	0.380	N.D. #
34)	L7 AR-1260-4	7.957	0.000	263118	0	2.448	N.D. #
35)	L7 AR-1260-5	8.266	0.000	44384	0	0.215	N.D. #
36)	L8 AR-1262-1	7.718	0.000	33395	0	0.283	N.D. #
37)	L8 AR-1262-2	8.266	0.000	44384	0	0.202	N.D. #
38)	L8 AR-1262-3	8.571	0.000	24399	0	0.160	N.D. #
39)	L8 AR-1262-4	8.665	0.000	176638	0	2.346	N.D. #
40)	L8 AR-1262-5	9.320	0.000	90519	0	1.098	N.D. #
41)	L9 AR-1268-1	8.571	0.000	24399	0	0.085	N.D. #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022823\
 Data File : PP056046.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Feb 2023 10:39
 Operator : YP\AJ
 Sample : PB151053BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 28 22:08:10 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 20 05:24:03 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	
42)	L9 AR-1268-2	8.672	0.000	159553	0	0.598	N.D.	#
43)	L9 AR-1268-3	8.898	0.000	149018	0	0.633	N.D.	#
44)	L9 AR-1268-4	9.320	0.000	90519	0	0.917	N.D.	#
45)	L9 AR-1268-5	9.733	8.326	72656	214680	0.097	0.532	#


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP022823\
Data File : PP056046.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Feb 2023 10:39
Operator : YP\AJ
Sample : PB151053BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 28 22:08:10 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP021723.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 20 05:24:03 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.323 min
Delta R.T.: -0.002 min
Response: 31464301
Conc: 19.53 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.566 min
Delta R.T.: -0.003 min
Response: 23302384
Conc: 15.96 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.072 min
Delta R.T.: -0.004 min
Response: 35102355
Conc: 17.21 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.570 min
Delta R.T.: -0.007 min
Response: 22975079
Conc: 20.25 ng/ml



#3 AR-1016-1


R.T.: 5.493 min
Delta R.T.: -0.003 min
Response: 8456
Conc: 0.14 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 4.652 min
Response: 0
Conc: N.D.




#4 AR-1016-2

R.T.: 5.523 min
Delta R.T.: 0.005 min
Response: 7750
Conc: 0.09 ng/ml



#4 AR-1016-2


R.T.: 0.000 min
Exp R.T. : 4.670 min
Response: 0
Conc: N.D.



#5 AR-1016-3


R.T.: 5.577 min
Delta R.T.: -0.004 min
Response: 62410
Conc: 1.12 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.848 min
Delta R.T.: 0.002 min
Response: 84536
Conc: 2.43 ng/ml




#6 AR-1016-4

R.T.: 5.677 min
Delta R.T.: -0.003 min
Response: 36181
Conc: 0.81 ng/ml



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T.: 4.889 min
Response: 0
Conc: N.D.



#7 AR-1016-5


R.T.: 5.977 min
Delta R.T.: 0.000 min
Response: 19166
Conc: 0.40 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 0.000 min
Exp R.T. : 5.102 min
Response: 0
Conc: N.D.




#8 AR-1221-1

R.T.: 4.537 min
Delta R.T.: 0.002 min
Response: 46018
Conc: 2.21 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 3.782 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.629 min
Delta R.T.: 0.008 min
Response: 524485
Conc: 32.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.864 min
Delta R.T.: -0.004 min
Response: 375009
Conc: 28.55 ng/ml




#10 AR-1221-3

R.T.: 4.694 min
Delta R.T.: -0.004 min
Response: 24482
Conc: 0.50 ng/ml



#10 AR-1221-3


R.T.: 3.953 min
Delta R.T.: 0.010 min
Response: 44626
Conc: 1.07 ng/ml



#11 AR-1232-1


R.T.: 4.694 min
Delta R.T.: -0.004 min
Response: 24482
Conc: 0.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.953 min
Delta R.T.: 0.011 min
Response: 44626
Conc: 1.30 ng/ml




#12 AR-1232-2

R.T.: 5.230 min
Delta R.T.: 0.001 min
Response: 9509
Conc: 0.46 ng/ml



#12 AR-1232-2


R.T.: 0.000 min
Exp R.T.: 4.670 min
Response: 0
Conc: N.D.



#13 AR-1232-3


R.T.: 5.523 min
Delta R.T.: 0.003 min
Response: 7750
Conc: 0.20 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.848 min
Delta R.T.: 0.002 min
Response: 84536
Conc: 5.56 ng/ml




#14 AR-1232-4

R.T.: 5.677 min
Delta R.T.: -0.004 min
Response: 36181
Conc: 1.85 ng/ml



#14 AR-1232-4


R.T.: 4.947 min
Delta R.T.: 0.016 min
Response: 230789
Conc: 15.22 ng/ml



#15 AR-1232-5


R.T.: 5.777 min
Delta R.T.: 0.005 min
Response: 40997
Conc: 2.56 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 0.000 min
Exp R.T. : 5.103 min
Response: 0
Conc: N.D.




#16 AR-1242-1

R.T.: 5.493 min
Delta R.T.: -0.004 min
Response: 8456
Conc: 0.18 ng/ml



#16 AR-1242-1


R.T.: 0.000 min
Exp R.T. : 4.652 min
Response: 0
Conc: N.D.



#17 AR-1242-2


R.T.: 5.523 min
Delta R.T.: 0.003 min
Response: 7750
Conc: 0.11 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 0.000 min
Exp R.T. : 4.670 min
Response: 0
Conc: N.D.




#18 AR-1242-3

R.T.: 5.577 min
Delta R.T.: -0.004 min
Response: 62410
Conc: 1.40 ng/ml



#18 AR-1242-3


R.T.: 4.848 min
Delta R.T.: 0.002 min
Response: 84536
Conc: 2.99 ng/ml



#19 AR-1242-4


R.T.: 5.677 min
Delta R.T.: -0.004 min
Response: 36181
Conc: 1.00 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.947 min
Delta R.T.: 0.017 min
Response: 230789
Conc: 7.48 ng/ml




#20 AR-1242-5

R.T.: 6.421 min
Delta R.T.: 0.000 min
Response: 47736
Conc: 1.14 ng/ml



#20 AR-1242-5


R.T.: 5.474 min
Delta R.T.: 0.019 min
Response: 844868
Conc: 25.36 ng/ml



#21 AR-1248-1


R.T.: 5.493 min
Delta R.T.: -0.004 min
Response: 8456
Conc: 0.23 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 0.000 min
Exp R.T. : 4.651 min
Response: 0
Conc: N.D.




#22 AR-1248-2

R.T.: 5.777 min
Delta R.T.: 0.006 min
Response: 40997
Conc: 0.71 ng/ml



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T. : 4.888 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 5.977 min
Delta R.T.: 0.001 min
Response: 19166
Conc: 0.29 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.947 min
Delta R.T.: 0.017 min
Response: 230789
Conc: 5.19 ng/ml




#24 AR-1248-4

R.T.: 6.389 min
Delta R.T.: 0.007 min
Response: 219221
Conc: 3.02 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T. : 5.101 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 6.421 min
Delta R.T.: 0.001 min
Response: 47736
Conc: 0.67 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.474 min
Delta R.T.: -0.021 min
Response: 844868
Conc: 19.17 ng/ml




#26 AR-1254-1

R.T.: 6.356 min
Delta R.T.: 0.001 min
Response: 750054
Conc: 10.20 ng/ml



#26 AR-1254-1


R.T.: 5.474 min
Delta R.T.: 0.020 min
Response: 844868
Conc: 11.48 ng/ml



#27 AR-1254-2


R.T.: 6.576 min
Delta R.T.: 0.002 min
Response: 59565
Conc: 0.52 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.605 min
Delta R.T.: 0.003 min
Response: 320437
Conc: 5.09 ng/ml




#28 AR-1254-3

R.T.: 6.941 min
Delta R.T.: 0.000 min
Response: 293103
Conc: 2.38 ng/ml



#28 AR-1254-3


R.T.: 6.033 min
Delta R.T.: 0.027 min
Response: 707267
Conc: 7.21 ng/ml



#29 AR-1254-4


R.T.: 7.217 min
Delta R.T.: -0.011 min
Response: 76793
Conc: 0.84 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 6.236 min
Response: 0
Conc: N.D.




#30 AR-1254-5

R.T.: 7.655 min
Delta R.T.: 0.005 min
Response: 504101
Conc: 4.80 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T. : 6.655 min
Response: 0
Conc: N.D.



#31 AR-1260-1


R.T.: 7.105 min
Delta R.T.: -0.003 min
Response: 121016
Conc: 1.17 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.132 min
Delta R.T.: -0.006 min
Response: 159379
Conc: 2.19 ng/ml




#32 AR-1260-2

R.T.: 7.370 min
Delta R.T.: 0.005 min
Response: 95282
Conc: 0.76 ng/ml



#32 AR-1260-2


R.T.: 6.307 min
Delta R.T.: -0.020 min
Response: 265105
Conc: 3.25 ng/ml



#33 AR-1260-3


R.T.: 7.718 min
Delta R.T.: -0.009 min
Response: 33395
Conc: 0.38 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 6.480 min
Response: 0
Conc: N.D.




#34 AR-1260-4

R.T.: 7.957 min
Delta R.T.: 0.004 min
Response: 263118
Conc: 2.45 ng/ml



#34 AR-1260-4


R.T.: 0.000 min
Exp R.T. : 6.954 min
Response: 0
Conc: N.D.



#35 AR-1260-5


R.T.: 8.266 min
Delta R.T.: -0.003 min
Response: 44384
Conc: 0.21 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 0.000 min
Exp R.T. : 7.197 min
Response: 0
Conc: N.D.




#36 AR-1262-1

R.T.: 7.718 min
Delta R.T.: -0.006 min
Response: 33395
Conc: 0.28 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T. : 6.745 min
Response: 0
Conc: N.D.



#37 AR-1262-2


R.T.: 8.266 min
Delta R.T.: -0.002 min
Response: 44384
Conc: 0.20 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 6.950 min
Response: 0
Conc: N.D.




#38 AR-1262-3

R.T.: 8.571 min
Delta R.T.: -0.007 min
Response: 24399
Conc: 0.16 ng/ml



#38 AR-1262-3


R.T.: 0.000 min
Exp R.T. : 7.478 min
Response: 0
Conc: N.D.



#39 AR-1262-4


R.T.: 8.665 min
Delta R.T.: -0.001 min
Response: 176638
Conc: 2.35 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 0.000 min
Exp R.T. : 7.542 min
Response: 0
Conc: N.D.




#40 AR-1262-5

R.T.: 9.320 min
Delta R.T.: -0.002 min
Response: 90519
Conc: 1.10 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T. : 8.038 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 8.571 min
Delta R.T.: -0.005 min
Response: 24399
Conc: 0.09 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 0.000 min
Exp R.T. : 7.478 min
Response: 0
Conc: N.D.




#42 AR-1268-2

R.T.: 8.672 min
Delta R.T.: 0.003 min
Response: 159553
Conc: 0.60 ng/ml



#42 AR-1268-2


R.T.: 0.000 min
Exp R.T. : 7.543 min
Response: 0
Conc: N.D.



#43 AR-1268-3


R.T.: 8.898 min
Delta R.T.: 0.002 min
Response: 149018
Conc: 0.63 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 7.748 min
Response: 0
Conc: N.D.




#44 AR-1268-4

R.T.: 9.320 min
Delta R.T.: 0.000 min
Response: 90519
Conc: 0.92 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T. : 8.037 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 9.733 min
Delta R.T.: 0.000 min
Response: 72656
Conc: 0.10 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.326 min
Delta R.T.: 0.002 min
Response: 214680
Conc: 0.53 ng/ml